

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080222\
 Data File : PP050078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2022 10:54
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 17:15:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.355	3.599	74566246	145.0E6	51.385	59.329m
2) SA Decachlor...	10.111	8.566	74230892	65401704	52.482	56.158
Target Compounds						
21) L5 AR-1248-1	5.532	4.671	17734255	28542055	503.449	507.594
22) L5 AR-1248-2	5.805	4.905	23835857	36374309	480.966	488.716
23) L5 AR-1248-3	6.012	4.946	26621352	37334143	480.677	488.939
24) L5 AR-1248-4	6.418	5.117	31432476	45200760	507.076	497.803
25) L5 AR-1248-5	6.456	5.508	29933498	40660322	508.476	487.853



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080222\
Data File : PP050078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2022 10:54
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 17:15:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.355 min
Delta R.T.: 0.002 min
Response: 74566246
Conc: 51.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.599 min
Delta R.T.: -0.003 min
Response: 145020027
Conc: 59.33 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.111 min
Delta R.T.: 0.004 min
Response: 74230892
Conc: 52.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.566 min
Delta R.T.: -0.006 min
Response: 65401704
Conc: 56.16 ng/ml



#21 AR-1248-1


R.T.: 5.532 min
Delta R.T.: 0.004 min
Response: 17734255
Conc: 503.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.671 min
Delta R.T.: -0.004 min
Response: 28542055
Conc: 507.59 ng/ml




#22 AR-1248-2

R.T.: 5.805 min
Delta R.T.: 0.004 min
Response: 23835857
Conc: 480.97 ng/ml



#22 AR-1248-2


R.T.: 4.905 min
Delta R.T.: -0.004 min
Response: 36374309
Conc: 488.72 ng/ml



#23 AR-1248-3


R.T.: 6.012 min
Delta R.T.: 0.004 min
Response: 26621352
Conc: 480.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.946 min
Delta R.T.: -0.004 min
Response: 37334143
Conc: 488.94 ng/ml




#24 AR-1248-4

R.T.: 6.418 min
Delta R.T.: 0.005 min
Response: 31432476
Conc: 507.08 ng/ml



#24 AR-1248-4


R.T.: 5.117 min
Delta R.T.: -0.005 min
Response: 45200760
Conc: 497.80 ng/ml



#25 AR-1248-5

R.T.: 6.456 min
Delta R.T.: 0.003 min
Response: 29933498
Conc: 508.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 40660322
Conc: 487.85 ng/ml