

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080322\
 Data File : PP050138.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2022 23:38
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 01:28:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.597	74792635	146.6E6	51.541	59.971
2) SA Decachlor...	10.093	8.558	66780978	57604073	47.215	49.463
Target Compounds						
3) L1 AR-1016-1	5.523	4.666	29533812	49109812	522.749	550.931m
4) L1 AR-1016-2	5.546	4.685	42554496	68906229	529.237	547.743
5) L1 AR-1016-3	5.607	4.860	25875612	35874280	528.363	545.914
6) L1 AR-1016-4	5.708	4.900	21747170	28116442	537.407	528.915
7) L1 AR-1016-5	6.002	5.112	20551922	34205889	502.127	497.492
31) L7 AR-1260-1	7.128	6.138	36041992	48586511	476.671	463.609
32) L7 AR-1260-2	7.387	6.325	43760891	57548935	462.273	478.276
33) L7 AR-1260-3	7.746	6.478	29356425	52405788	468.800	463.138
34) L7 AR-1260-4	7.974	6.948	35639567	35462292	460.610	463.608
35) L7 AR-1260-5	8.294	7.189	67157176	79419160	436.794	464.442



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080322\
Data File : PP050138.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2022 23:38
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 01:28:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: -0.005 min
Response: 74792635
Conc: 51.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: -0.005 min
Response: 146588598
Conc: 59.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.093 min
Delta R.T.: -0.014 min
Response: 66780978
Conc: 47.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.558 min
Delta R.T.: -0.015 min
Response: 57604073
Conc: 49.46 ng/ml



#3 AR-1016-1


R.T.: 5.523 min
Delta R.T.: -0.008 min
Response: 29533812
Conc: 522.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.666 min
Delta R.T.: -0.010 min
Response: 49109812
Conc: 550.93 ng/ml m




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: -0.007 min
Response: 42554496
Conc: 529.24 ng/ml



#4 AR-1016-2


R.T.: 4.685 min
Delta R.T.: -0.010 min
Response: 68906229
Conc: 547.74 ng/ml



#5 AR-1016-3


R.T.: 5.607 min
Delta R.T.: -0.008 min
Response: 25875612
Conc: 528.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.860 min
Delta R.T.: -0.011 min
Response: 35874280
Conc: 545.91 ng/ml




#6 AR-1016-4

R.T.: 5.708 min
Delta R.T.: -0.007 min
Response: 21747170
Conc: 537.41 ng/ml



#6 AR-1016-4


R.T.: 4.900 min
Delta R.T.: -0.011 min
Response: 28116442
Conc: 528.91 ng/ml



#7 AR-1016-5


R.T.: 6.002 min
Delta R.T.: -0.008 min
Response: 20551922
Conc: 502.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.112 min
Delta R.T.: -0.011 min
Response: 34205889
Conc: 497.49 ng/ml




#31 AR-1260-1

R.T.: 7.128 min
Delta R.T.: -0.009 min
Response: 36041992
Conc: 476.67 ng/ml



#31 AR-1260-1


R.T.: 6.138 min
Delta R.T.: -0.013 min
Response: 48586511
Conc: 463.61 ng/ml



#32 AR-1260-2


R.T.: 7.387 min
Delta R.T.: -0.008 min
Response: 43760891
Conc: 462.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.325 min
Delta R.T.: -0.012 min
Response: 57548935
Conc: 478.28 ng/ml




#33 AR-1260-3

R.T.: 7.746 min
Delta R.T.: -0.008 min
Response: 29356425
Conc: 468.80 ng/ml



#33 AR-1260-3


R.T.: 6.478 min
Delta R.T.: -0.013 min
Response: 52405788
Conc: 463.14 ng/ml



#34 AR-1260-4


R.T.: 7.974 min
Delta R.T.: -0.009 min
Response: 35639567
Conc: 460.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.948 min
Delta R.T.: -0.013 min
Response: 35462292
Conc: 463.61 ng/ml



#35 AR-1260-5

R.T.: 8.294 min
Delta R.T.: -0.009 min
Response: 67157176
Conc: 436.79 ng/ml



#35 AR-1260-5

R.T.: 7.189 min
Delta R.T.: -0.013 min
Response: 79419160
Conc: 464.44 ng/ml