

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080322\
 Data File : PP050152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2022 04:39
 Operator : YP\AJ
 Sample : N4017-08
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB05

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 06:39:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.598	34103186	69011861	23.501	28.233
2) SA Decachlor...	10.095	8.559	31475197	27971395	22.253	24.018

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080322\
Data File : PP050152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2022 04:39
Operator : YP\AJ
Sample : N4017-08
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB05

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 06:39:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: -0.005 min
Response: 34103186
Conc: 23.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB05




#1 Tetrachloro-m-xylene

R.T.: 3.598 min
Delta R.T.: -0.004 min
Response: 69011861
Conc: 28.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: -0.012 min
Response: 31475197
Conc: 22.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.559 min
Delta R.T.: -0.014 min
Response: 27971395
Conc: 24.02 ng/ml