

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080422\
 Data File : PP050186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2022 21:16
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 02:42:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.599	76974588	156.7E6	53.044	64.114
2) SA Decachlor...	10.096	8.560	62466391	51855090	44.164	44.526
Target Compounds						
21) L5 AR-1248-1	5.525	4.668	17948891	29790787	509.542	529.802
22) L5 AR-1248-2	5.798	4.902	23537131	37705372	474.938	506.600
23) L5 AR-1248-3	6.004	4.944	25925570	38285256	468.114	501.395
24) L5 AR-1248-4	6.410	5.114	29787933	43560914	480.546	479.743
25) L5 AR-1248-5	6.449	5.504	28164526	38550477	478.427	462.538



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080422\
Data File : PP050186.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2022 21:16
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 02:42:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: -0.003 min
Response: 76974588
Conc: 53.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.599 min
Delta R.T.: -0.003 min
Response: 156714900
Conc: 64.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.096 min
Delta R.T.: -0.011 min
Response: 62466391
Conc: 44.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.560 min
Delta R.T.: -0.013 min
Response: 51855090
Conc: 44.53 ng/ml



#21 AR-1248-1


R.T.: 5.525 min
Delta R.T.: -0.003 min
Response: 17948891
Conc: 509.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.668 min
Delta R.T.: -0.007 min
Response: 29790787
Conc: 529.80 ng/ml




#22 AR-1248-2

R.T.: 5.798 min
Delta R.T.: -0.004 min
Response: 23537131
Conc: 474.94 ng/ml



#22 AR-1248-2


R.T.: 4.902 min
Delta R.T.: -0.007 min
Response: 37705372
Conc: 506.60 ng/ml



#23 AR-1248-3


R.T.: 6.004 min
Delta R.T.: -0.004 min
Response: 25925570
Conc: 468.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.944 min
Delta R.T.: -0.007 min
Response: 38285256
Conc: 501.40 ng/ml




#24 AR-1248-4

R.T.: 6.410 min
Delta R.T.: -0.004 min
Response: 29787933
Conc: 480.55 ng/ml



#24 AR-1248-4


R.T.: 5.114 min
Delta R.T.: -0.008 min
Response: 43560914
Conc: 479.74 ng/ml



#25 AR-1248-5

R.T.: 6.449 min
Delta R.T.: -0.004 min
Response: 28164526
Conc: 478.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.504 min
Delta R.T.: -0.008 min
Response: 38550477
Conc: 462.54 ng/ml