

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080423\
 Data File : PP059128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2023 14:04
 Operator : YP\AJ
 Sample : 03891-01MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 06:13:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 01 05:06:30 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.420	3.655	34213835	48221217	21.203	19.942
2) SA Decachlor...	10.217	8.694	18434106	28097872	19.667	20.875
Target Compounds						
3) L1 AR-1016-1	5.597	4.749	24101975	34220867	511.501	508.179
4) L1 AR-1016-2	5.619	4.768	34821493	48909768	513.128	513.874
5) L1 AR-1016-3	5.682	4.945	22289155	26468001	515.528	518.559
6) L1 AR-1016-4	5.781	4.987	18237722	23398618	516.837	521.198
7) L1 AR-1016-5	6.078	5.201	19249624	27666712	516.016	496.572
31) L7 AR-1260-1	7.210	6.240	34948049	51079089	569.551	505.098
32) L7 AR-1260-2	7.467	6.429	36332684	56678967	562.264	497.104
33) L7 AR-1260-3	7.828	6.583	23670951	54164992	491.434	496.457
34) L7 AR-1260-4	8.055	7.057	26959601	38503215	497.080	449.459
35) L7 AR-1260-5	8.376	7.300	42858504	79746787	433.814	456.496



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080423\
Data File : PP059128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2023 14:04
Operator : YP\AJ
Sample : 03891-01MS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 06:13:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 01 05:06:30 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.420 min
Delta R.T.: -0.002 min
Response: 34213835
Conc: 21.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#1 Tetrachloro-m-xylene

R.T.: 3.655 min
Delta R.T.: -0.002 min
Response: 48221217
Conc: 19.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.001 min
Response: 18434106
Conc: 19.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.694 min
Delta R.T.: -0.004 min
Response: 28097872
Conc: 20.87 ng/ml



#3 AR-1016-1


R.T.: 5.597 min
Delta R.T.: -0.002 min
Response: 24101975
Conc: 511.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#3 AR-1016-1

R.T.: 4.749 min
Delta R.T.: -0.003 min
Response: 34220867
Conc: 508.18 ng/ml




#4 AR-1016-2

R.T.: 5.619 min
Delta R.T.: -0.002 min
Response: 34821493
Conc: 513.13 ng/ml



#4 AR-1016-2


R.T.: 4.768 min
Delta R.T.: -0.003 min
Response: 48909768
Conc: 513.87 ng/ml



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: -0.002 min
Response: 22289155
Conc: 515.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#5 AR-1016-3

R.T.: 4.945 min
Delta R.T.: -0.003 min
Response: 26468001
Conc: 518.56 ng/ml




#6 AR-1016-4

R.T.: 5.781 min
Delta R.T.: -0.002 min
Response: 18237722
Conc: 516.84 ng/ml



#6 AR-1016-4


R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 23398618
Conc: 521.20 ng/ml



#7 AR-1016-5


R.T.: 6.078 min
Delta R.T.: -0.002 min
Response: 19249624
Conc: 516.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#7 AR-1016-5

R.T.: 5.201 min
Delta R.T.: -0.003 min
Response: 27666712
Conc: 496.57 ng/ml




#31 AR-1260-1

R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 34948049
Conc: 569.55 ng/ml



#31 AR-1260-1


R.T.: 6.240 min
Delta R.T.: -0.004 min
Response: 51079089
Conc: 505.10 ng/ml



#32 AR-1260-2


R.T.: 7.467 min
Delta R.T.: -0.002 min
Response: 36332684
Conc: 562.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#32 AR-1260-2

R.T.: 6.429 min
Delta R.T.: -0.003 min
Response: 56678967
Conc: 497.10 ng/ml




#33 AR-1260-3

R.T.: 7.828 min
Delta R.T.: -0.002 min
Response: 23670951
Conc: 491.43 ng/ml



#33 AR-1260-3


R.T.: 6.583 min
Delta R.T.: -0.003 min
Response: 54164992
Conc: 496.46 ng/ml



#34 AR-1260-4


R.T.: 8.055 min
Delta R.T.: 0.000 min
Response: 26959601
Conc: 497.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-1MS




#34 AR-1260-4

R.T.: 7.057 min
Delta R.T.: -0.003 min
Response: 38503215
Conc: 449.46 ng/ml



#35 AR-1260-5

R.T.: 8.376 min
Delta R.T.: 0.000 min
Response: 42858504
Conc: 433.81 ng/ml



#35 AR-1260-5

R.T.: 7.300 min
Delta R.T.: -0.003 min
Response: 79746787
Conc: 456.50 ng/ml