

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080521\
 Data File : PP038021.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Aug 2021 10:55
 Operator : AJ\MA
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 12:00:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 05 11:58:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.894	3.879	4114962	2485469	95.491	96.641
2) SA Decachlor...	10.888	9.151	2157604	2374658	94.399	92.303
Target Compounds						
26) L6 AR-1254-1	7.121	5.995	1116623	1194349	901.681	919.337
27) L6 AR-1254-2	7.351	6.155	1649981	1030045	904.174	909.728
28) L6 AR-1254-3	7.733	6.573	1763405	1722537	915.816	931.865
29) L6 AR-1254-4	8.029	6.814	1278666	1108467	914.611	926.578
30) L6 AR-1254-5	8.458	7.242	1322708	1539059	921.372	934.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080521\
Data File : PP038021.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Aug 2021 10:55
Operator : AJ\MA
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 12:00:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 05 11:58:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.894 min
Delta R.T.: -0.002 min
Response: 4114962
Conc: 95.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.879 min
Delta R.T.: -0.002 min
Response: 2485469
Conc: 96.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.888 min
Delta R.T.: -0.002 min
Response: 2157604
Conc: 94.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.151 min
Delta R.T.: 0.000 min
Response: 2374658
Conc: 92.30 ng/ml



#26 AR-1254-1


R.T.: 7.121 min
Delta R.T.: -0.002 min
Response: 1116623
Conc: 901.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.995 min
Delta R.T.: -0.001 min
Response: 1194349
Conc: 919.34 ng/ml




#27 AR-1254-2

R.T.: 7.351 min
Delta R.T.: -0.002 min
Response: 1649981
Conc: 904.17 ng/ml



#27 AR-1254-2


R.T.: 6.155 min
Delta R.T.: -0.002 min
Response: 1030045
Conc: 909.73 ng/ml



#28 AR-1254-3


R.T.: 7.733 min
Delta R.T.: -0.002 min
Response: 1763405
Conc: 915.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.573 min
Delta R.T.: -0.001 min
Response: 1722537
Conc: 931.87 ng/ml




#29 AR-1254-4

R.T.: 8.029 min
Delta R.T.: -0.002 min
Response: 1278666
Conc: 914.61 ng/ml



#29 AR-1254-4


R.T.: 6.814 min
Delta R.T.: 0.000 min
Response: 1108467
Conc: 926.58 ng/ml



#30 AR-1254-5

R.T.: 8.458 min
Delta R.T.: -0.002 min
Response: 1322708
Conc: 921.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 7.242 min
Delta R.T.: -0.002 min
Response: 1539059
Conc: 934.66 ng/ml