

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080624\
 Data File : PP066596.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2024 11:41
 Operator : YP\AJ
 Sample : P3390-07
 Misc : AR1248 LOD 25 PPB
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 13:59:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 06 04:00:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.781	4.019	30973822	33629859	16.378	15.969
2) SA Decachlor...	10.777	9.194	40133969	24265678	19.209	18.254
Target Compounds						
21) L5 AR-1248-1	5.961	5.132	1075334	813946	22.243	20.360
22) L5 AR-1248-2	6.235	5.372	1774729	1085699	25.038	20.501
23) L5 AR-1248-3	6.443	5.416	1784608	934460	23.093	17.057 #
24) L5 AR-1248-4	6.848	5.593	2067683	1161770	23.258	18.513
25) L5 AR-1248-5	6.889	5.992	2105937	1074780	23.841	18.563



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080624\
Data File : PP066596.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2024 11:41
Operator : YP\AJ
Sample : P3390-07
Misc : AR1248 LOD 25 PPB
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 13:59:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080524.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 06 04:00:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.781 min
Delta R.T.: 0.002 min
Response: 30973822
Conc: 16.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024




#1 Tetrachloro-m-xylene

R.T.: 4.019 min
Delta R.T.: 0.001 min
Response: 33629859
Conc: 15.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.777 min
Delta R.T.: -0.001 min
Response: 40133969
Conc: 19.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.194 min
Delta R.T.: 0.000 min
Response: 24265678
Conc: 18.25 ng/ml



#21 AR-1248-1


R.T.: 5.961 min
Delta R.T.: 0.000 min
Response: 1075334
Conc: 22.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024




#21 AR-1248-1

R.T.: 5.132 min
Delta R.T.: 0.000 min
Response: 813946
Conc: 20.36 ng/ml




#22 AR-1248-2

R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 1774729
Conc: 25.04 ng/ml



#22 AR-1248-2


R.T.: 5.372 min
Delta R.T.: 0.000 min
Response: 1085699
Conc: 20.50 ng/ml



#23 AR-1248-3


R.T.: 6.443 min
Delta R.T.: 0.000 min
Response: 1784608
Conc: 23.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024




#23 AR-1248-3

R.T.: 5.416 min
Delta R.T.: -0.001 min
Response: 934460
Conc: 17.06 ng/ml




#24 AR-1248-4

R.T.: 6.848 min
Delta R.T.: 0.000 min
Response: 2067683
Conc: 23.26 ng/ml



#24 AR-1248-4


R.T.: 5.593 min
Delta R.T.: -0.001 min
Response: 1161770
Conc: 18.51 ng/ml



#25 AR-1248-5

R.T.: 6.889 min
Delta R.T.: 0.000 min
Response: 2105937
Conc: 23.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024



#25 AR-1248-5

R.T.: 5.992 min
Delta R.T.: -0.001 min
Response: 1074780
Conc: 18.56 ng/ml