

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080718\
 Data File : PP026997.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Aug 2018 00:06
 Operator : EI\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 01:08:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073018.M
 Quant Title : 8080.M
 QLast Update : Tue Jul 31 01:31:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.044	12.156	125.0E6	192.0E6	523.317	474.086
Target Compounds							
12) T	2,4,5-T	15.566	0.000	22424956	0	8.788	N.D. #
15) T	Picloram	17.598	0.000	16093240	0	3.897	N.D. #
16) T	DCPA	18.256	0.000	5680541	0	1.416	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080718\
Data File : PP026997.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Aug 2018 00:06
Operator : EI\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 01:08:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073018.M
Quant Title : 8080.M
QLast Update : Tue Jul 31 01:31:14 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 12.044 min
Delta R.T.: 0.000 min
Response: 124980109
Conc: 523.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 12.156 min
Delta R.T.: 0.000 min
Response: 191972521
Conc: 474.09 ng/ml




#12 2,4,5-T

R.T.: 15.566 min
Delta R.T.: -0.013 min
Response: 22424956
Conc: 8.79 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 15.886 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 17.598 min
Delta R.T.: 0.011 min
Response: 16093240
Conc: 3.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#15 Picloram

R.T.: 0.000 min
Exp R.T.: 18.255 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 18.256 min
Delta R.T.: 0.060 min
Response: 5680541
Conc: 1.42 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 18.143 min
Response: 0
Conc: N.D.