

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080723\
 Data File : PP059181.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Aug 2023 18:50
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 01:32:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 01:32:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.423	3.658	176.9E6	283.6E6	97.171	97.354
2)	SA Decachlor...	10.221	8.694	107.3E6	172.6E6	96.460	97.552
Target Compounds							
26)	L6 AR-1254-1	6.460	5.556	73559007	144.1E6	941.736	959.617
27)	L6 AR-1254-2	6.679	5.705	112.0E6	125.8E6	948.993	954.784
28)	L6 AR-1254-3	7.047	6.110	118.1E6	204.7E6	960.479	968.712
29)	L6 AR-1254-4	7.333	6.340	76626428	112.5E6	956.581	969.747
30)	L6 AR-1254-5	7.755	6.760	83418974	175.9E6	961.752	973.305

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080723\
Data File : PP059181.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Aug 2023 18:50
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 01:32:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 01:32:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

