

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080723\
 Data File : PP059196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Aug 2023 22:55
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080723AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 02:05:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 08 01:54:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.656	90838457	143.9E6	50.334	51.143
2) SA Decachlor...	10.217	8.691	106.6E6	170.9E6	52.082	52.743
Target Compounds						
41) L9 AR-1268-1	8.691	7.582	103.5E6	184.5E6	517.893	517.605
42) L9 AR-1268-2	8.785	7.647	89170934	165.7E6	514.846	521.174
43) L9 AR-1268-3	9.017	7.855	82019478	150.2E6	520.398	521.362
44) L9 AR-1268-4	9.448	8.143	28448619	49531508	528.780	532.340
45) L9 AR-1268-5	9.871	8.435	197.1E6	358.7E6	528.218	529.806

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080723\
Data File : PP059196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Aug 2023 22:55
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP080723AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 02:05:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 08 01:54:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

