

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080822\
 Data File : PP050294.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2022 09:23
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 02:53:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Aug 07 23:52:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
6)	L1	AR-1016-4	0.000	4.905	0	98418 N.D. 1.689 #
14)	L3	AR-1232-4	0.000	4.919f	0	64199 N.D. 2.179 #
19)	L4	AR-1242-4	0.000	4.919f	0	64199 N.D. 1.097 #
20)	L4	AR-1242-5	0.000	5.483f	0	405728 N.D. 5.837 #
22)	L5	AR-1248-2	0.000	4.905	0	98418 N.D. 1.139 #
23)	L5	AR-1248-3	0.000	4.919f	0	64199 N.D. 0.708 #
25)	L5	AR-1248-5	0.000	5.483f	0	405728 N.D. 4.086 #
26)	L6	AR-1254-1	0.000	5.483f	0	405728 N.D. 3.070 #
29)	L6	AR-1254-4	7.261	0.000	7018518	0 77.018 N.D. #
32)	L7	AR-1260-2	0.000	6.310f	0	326660 N.D. 2.432 #
33)	L7	AR-1260-3	7.755	0.000	19505958	0 229.059 N.D. #
36)	L8	AR-1262-1	7.755	0.000	19505958	0 162.879 N.D. #
45)	L9	AR-1268-5	9.780f	0.000	697081	0 1.045 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080822\
Data File : PP050294.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2022 09:23
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 02:53:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Aug 07 23:52:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.346 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.593 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.090 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.555 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 5.707 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#6 AR-1016-4

R.T.: 4.905 min
Delta R.T.: 0.005 min
Response: 98418
Conc: 1.69 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T.: 5.708 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 4.919 min
Delta R.T.: -0.022 min
Response: 64199
Conc: 2.18 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 5.708 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.919 min
Delta R.T.: -0.022 min
Response: 64199
Conc: 1.10 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.446 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.483 min
Delta R.T.: 0.022 min
Response: 405728
Conc: 5.84 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.794 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#22 AR-1248-2

R.T.: 4.905 min
Delta R.T.: 0.006 min
Response: 98418
Conc: 1.14 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T.: 6.000 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 4.919 min
Delta R.T.: -0.021 min
Response: 64199
Conc: 0.71 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.445 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.483 min
Delta R.T.: -0.018 min
Response: 405728
Conc: 4.09 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.379 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.483 min
Delta R.T.: 0.022 min
Response: 405728
Conc: 3.07 ng/ml



#29 AR-1254-4


R.T.: 7.261 min
Delta R.T.: 0.006 min
Response: 7018518
Conc: 77.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.236 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.386 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.310 min
Delta R.T.: -0.015 min
Response: 326660
Conc: 2.43 ng/ml



#33 AR-1260-3


R.T.: 7.755 min
Delta R.T.: 0.010 min
Response: 19505958
Conc: 229.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.478 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.755 min
Delta R.T.: 0.011 min
Response: 19505958
Conc: 162.88 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.688 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.780 min
Delta R.T.: 0.022 min
Response: 697081
Conc: 1.05 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.309 min
Response: 0
Conc: N.D.