

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080822\
 Data File : PP050300.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2022 11:03
 Operator : YP\AJ
 Sample : N3980-08
 Misc : AR1242 LOQ 50 PPB
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOQ-WATER-02-QT3-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 02:57:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Aug 07 23:52:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.345	3.592	31482394	60927718	19.558	19.452
2)	SA Decachlor...	10.094	8.558	34548992	28811467	19.620	18.720
Target Compounds							
16)	L4 AR-1242-1	5.523	4.665	2329965	4019112	45.972	48.482
17)	L4 AR-1242-2	5.546	4.684	3370331	5566314	46.655	47.665
18)	L4 AR-1242-3	5.608	4.858	1995974	2865435	45.134	46.729
19)	L4 AR-1242-4	5.709	4.941	1696392	2787033	45.523	47.609
20)	L4 AR-1242-5	6.447	5.460	2041357	3568253	48.372	51.335

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080822\
Data File : PP050300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2022 11:03
Operator : YP\AJ
Sample : N3980-08
Misc : AR1242 LOQ 50 PPB
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 02:57:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Aug 07 23:52:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

