

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080823\
 Data File : PP059242.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 06:14
 Operator : YP\AJ
 Sample : 03572-01
 Misc : AR1254 LOD 40 PPB
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 06:55:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 02:30:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.414	3.647	24070495	40728343	19.617	20.101
2)	SA Decachlor...	10.207	8.683	19008204	33562636	24.905	23.857
Target Compounds							
3)	L1 AR-1016-1	5.593	4.741	115286	215774	2.934	3.364
4)	L1 AR-1016-2	5.613	4.759	84912	132642	1.498	1.460
5)	L1 AR-1016-3	5.665	4.940	136038	279610	3.792	5.819 #
6)	L1 AR-1016-4	5.772	4.978	30958	2096507	1.069	50.774 #
7)	L1 AR-1016-5	6.071	5.191	705668	1240786	23.317	23.846
8)	L2 AR-1221-1	4.624	3.863	70611	41456	4.482	1.527 #
9)	L2 AR-1221-2	4.724	3.949	640831	730966	54.267	36.714 #
10)	L2 AR-1221-3	4.790	4.027	402231	92396	11.423	1.552 #
11)	L3 AR-1232-1	4.790	4.027	402231	92396	13.683	1.947 #
12)	L3 AR-1232-2	5.320	4.759	49122	132642	3.151	3.127
13)	L3 AR-1232-3	5.613	4.940	84912	279610	3.216	12.602 #
14)	L3 AR-1232-4	5.772	5.019	30958	828530	2.339	39.240 #
15)	L3 AR-1232-5	5.866	5.191	1097480	1240786	96.147	54.129 #
16)	L4 AR-1242-1	5.593	4.741	115286	215774	3.368	3.856
17)	L4 AR-1242-2	5.613	4.759	84912	132642	1.730	1.663
18)	L4 AR-1242-3	5.665	4.940	136038	279610	4.330	6.561 #
19)	L4 AR-1242-4	5.772	5.019	30958	828530	1.207	18.974 #
20)	L4 AR-1242-5	6.518	5.544	617641	4509527	23.908	90.785 #
21)	L5 AR-1248-1	5.593	4.741	115286	215774	4.542	5.245
22)	L5 AR-1248-2	5.866	4.978	1097480	2096507	28.355	34.816
23)	L5 AR-1248-3	6.071	5.019	705668	828530	16.704	13.448
24)	L5 AR-1248-4	6.476	5.191	900576	1240786	21.630	17.234
25)	L5 AR-1248-5	6.518	5.586	617641	660293	15.141	10.007 #
26)	L6 AR-1254-1	6.451	5.544	2055332	4509527	43.997	44.510
27)	L6 AR-1254-2	6.671	5.694	3099848	4008783	45.051	45.411
28)	L6 AR-1254-3	7.038	6.098	3331699	5917918	48.271	42.807
29)	L6 AR-1254-4	7.325	6.329	2017736	3391713	44.823	42.196
30)	L6 AR-1254-5	7.746	6.748	2078747	5073134	44.241	42.422
31)	L7 AR-1260-1	7.203	6.229	1525440	2984408	27.883	30.134
32)	L7 AR-1260-2	7.461	6.419	1395208	2533918	23.434	21.899
33)	L7 AR-1260-3	7.834	6.571	23156	3027485	0.588	27.217 #
34)	L7 AR-1260-4	8.034	7.049	529725	95343	11.927	1.100 #
35)	L7 AR-1260-5	8.368	7.289	257472	804018	3.220	4.362 #
36)	L8 AR-1262-1	7.834	6.844	23156	61354	0.364	1.066 #
37)	L8 AR-1262-2	8.368	7.049	257472	95343	2.804	0.808 #
38)	L8 AR-1262-3	8.701	7.574	179934	11605	2.796	0.136 #
39)	L8 AR-1262-4	8.773	7.635	20017	660240	0.603	4.371 #
40)	L8 AR-1262-5	9.431	8.138	18346	135035	0.615	2.117 #
41)	L9 AR-1268-1	8.701f	7.574	179934	11605	1.584	0.048 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080823\
 Data File : PP059242.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 06:14
 Operator : YP\AJ
 Sample : 03572-01
 Misc : AR1254 LOD 40 PPB
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 06:55:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 02:30:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.784	7.635	25528	660240	0.259	3.015 #
43) L9	AR-1268-3	9.005	7.851	52561	240212	0.567	1.228 #
44) L9	AR-1268-4	9.431	8.138	18346	135035	0.575	1.940 #
45) L9	AR-1268-5	9.864	8.428	66591	205157	0.245	0.396 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080823\
Data File : PP059242.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2023 06:14
Operator : YP\AJ
Sample : 03572-01
Misc : AR1254 LOD 40 PPB
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 06:55:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 09 02:30:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.002 min
Response: 24070495
Conc: 19.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.002 min
Response: 40728343
Conc: 20.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.005 min
Response: 19008204
Conc: 24.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: -0.006 min
Response: 33562636
Conc: 23.86 ng/ml



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 115286
Conc: 2.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.741 min
Delta R.T.: -0.002 min
Response: 215774
Conc: 3.36 ng/ml




#4 AR-1016-2

R.T.: 5.613 min
Delta R.T.: -0.003 min
Response: 84912
Conc: 1.50 ng/ml



#4 AR-1016-2


R.T.: 4.759 min
Delta R.T.: -0.003 min
Response: 132642
Conc: 1.46 ng/ml



#5 AR-1016-3


R.T.: 5.665 min
Delta R.T.: -0.014 min
Response: 136038
Conc: 3.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.940 min
Delta R.T.: 0.002 min
Response: 279610
Conc: 5.82 ng/ml




#6 AR-1016-4

R.T.: 5.772 min
Delta R.T.: -0.006 min
Response: 30958
Conc: 1.07 ng/ml



#6 AR-1016-4


R.T.: 4.978 min
Delta R.T.: -0.003 min
Response: 2096507
Conc: 50.77 ng/ml



#7 AR-1016-5


R.T.: 6.071 min
Delta R.T.: -0.004 min
Response: 705668
Conc: 23.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.191 min
Delta R.T.: -0.004 min
Response: 1240786
Conc: 23.85 ng/ml




#8 AR-1221-1

R.T.: 4.624 min
Delta R.T.: -0.002 min
Response: 70611
Conc: 4.48 ng/ml



#8 AR-1221-1


R.T.: 3.863 min
Delta R.T.: -0.002 min
Response: 41456
Conc: 1.53 ng/ml



#9 AR-1221-2


R.T.: 4.724 min
Delta R.T.: 0.010 min
Response: 640831
Conc: 54.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.949 min
Delta R.T.: -0.002 min
Response: 730966
Conc: 36.71 ng/ml




#10 AR-1221-3

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 402231
Conc: 11.42 ng/ml



#10 AR-1221-3


R.T.: 4.027 min
Delta R.T.: 0.000 min
Response: 92396
Conc: 1.55 ng/ml



#11 AR-1232-1


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 402231
Conc: 13.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.027 min
Delta R.T.: 0.000 min
Response: 92396
Conc: 1.95 ng/ml




#12 AR-1232-2

R.T.: 5.320 min
Delta R.T.: -0.004 min
Response: 49122
Conc: 3.15 ng/ml



#12 AR-1232-2


R.T.: 4.759 min
Delta R.T.: -0.003 min
Response: 132642
Conc: 3.13 ng/ml



#13 AR-1232-3


R.T.: 5.613 min
Delta R.T.: -0.003 min
Response: 84912
Conc: 3.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.940 min
Delta R.T.: 0.003 min
Response: 279610
Conc: 12.60 ng/ml




#14 AR-1232-4

R.T.: 5.772 min
Delta R.T.: -0.006 min
Response: 30958
Conc: 2.34 ng/ml



#14 AR-1232-4


R.T.: 5.019 min
Delta R.T.: -0.003 min
Response: 828530
Conc: 39.24 ng/ml



#15 AR-1232-5


R.T.: 5.866 min
Delta R.T.: -0.003 min
Response: 1097480
Conc: 96.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.191 min
Delta R.T.: -0.003 min
Response: 1240786
Conc: 54.13 ng/ml




#16 AR-1242-1

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 115286
Conc: 3.37 ng/ml



#16 AR-1242-1


R.T.: 4.741 min
Delta R.T.: -0.001 min
Response: 215774
Conc: 3.86 ng/ml



#17 AR-1242-2


R.T.: 5.613 min
Delta R.T.: -0.003 min
Response: 84912
Conc: 1.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.759 min
Delta R.T.: -0.002 min
Response: 132642
Conc: 1.66 ng/ml




#18 AR-1242-3

R.T.: 5.665 min
Delta R.T.: -0.013 min
Response: 136038
Conc: 4.33 ng/ml



#18 AR-1242-3


R.T.: 4.940 min
Delta R.T.: 0.003 min
Response: 279610
Conc: 6.56 ng/ml



#19 AR-1242-4


R.T.: 5.772 min
Delta R.T.: -0.005 min
Response: 30958
Conc: 1.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.019 min
Delta R.T.: -0.002 min
Response: 828530
Conc: 18.97 ng/ml




#20 AR-1242-5

R.T.: 6.518 min
Delta R.T.: 0.000 min
Response: 617641
Conc: 23.91 ng/ml



#20 AR-1242-5


R.T.: 5.544 min
Delta R.T.: -0.003 min
Response: 4509527
Conc: 90.78 ng/ml



#21 AR-1248-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 115286
Conc: 4.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.741 min
Delta R.T.: -0.001 min
Response: 215774
Conc: 5.25 ng/ml




#22 AR-1248-2

R.T.: 5.866 min
Delta R.T.: -0.003 min
Response: 1097480
Conc: 28.35 ng/ml



#22 AR-1248-2


R.T.: 4.978 min
Delta R.T.: -0.003 min
Response: 2096507
Conc: 34.82 ng/ml



#23 AR-1248-3


R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 705668
Conc: 16.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.019 min
Delta R.T.: -0.002 min
Response: 828530
Conc: 13.45 ng/ml




#24 AR-1248-4

R.T.: 6.476 min
Delta R.T.: -0.004 min
Response: 900576
Conc: 21.63 ng/ml



#24 AR-1248-4


R.T.: 5.191 min
Delta R.T.: -0.003 min
Response: 1240786
Conc: 17.23 ng/ml



#25 AR-1248-5


R.T.: 6.518 min
Delta R.T.: 0.000 min
Response: 617641
Conc: 15.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.586 min
Delta R.T.: -0.004 min
Response: 660293
Conc: 10.01 ng/ml




#26 AR-1254-1

R.T.: 6.451 min
Delta R.T.: -0.002 min
Response: 2055332
Conc: 44.00 ng/ml



#26 AR-1254-1


R.T.: 5.544 min
Delta R.T.: -0.004 min
Response: 4509527
Conc: 44.51 ng/ml



#27 AR-1254-2


R.T.: 6.671 min
Delta R.T.: -0.003 min
Response: 3099848
Conc: 45.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.694 min
Delta R.T.: -0.003 min
Response: 4008783
Conc: 45.41 ng/ml




#28 AR-1254-3

R.T.: 7.038 min
Delta R.T.: -0.002 min
Response: 3331699
Conc: 48.27 ng/ml



#28 AR-1254-3


R.T.: 6.098 min
Delta R.T.: -0.004 min
Response: 5917918
Conc: 42.81 ng/ml



#29 AR-1254-4


R.T.: 7.325 min
Delta R.T.: -0.003 min
Response: 2017736
Conc: 44.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.329 min
Delta R.T.: -0.004 min
Response: 3391713
Conc: 42.20 ng/ml




#30 AR-1254-5

R.T.: 7.746 min
Delta R.T.: -0.002 min
Response: 2078747
Conc: 44.24 ng/ml



#30 AR-1254-5


R.T.: 6.748 min
Delta R.T.: -0.004 min
Response: 5073134
Conc: 42.42 ng/ml



#31 AR-1260-1


R.T.: 7.203 min
Delta R.T.: -0.003 min
Response: 1525440
Conc: 27.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.229 min
Delta R.T.: -0.004 min
Response: 2984408
Conc: 30.13 ng/ml




#32 AR-1260-2

R.T.: 7.461 min
Delta R.T.: -0.003 min
Response: 1395208
Conc: 23.43 ng/ml



#32 AR-1260-2


R.T.: 6.419 min
Delta R.T.: -0.004 min
Response: 2533918
Conc: 21.90 ng/ml



#33 AR-1260-3


R.T.: 7.834 min
Delta R.T.: 0.008 min
Response: 23156
Conc: 0.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.571 min
Delta R.T.: -0.005 min
Response: 3027485
Conc: 27.22 ng/ml




#34 AR-1260-4

R.T.: 8.034 min
Delta R.T.: -0.017 min
Response: 529725
Conc: 11.93 ng/ml



#34 AR-1260-4


R.T.: 7.049 min
Delta R.T.: 0.000 min
Response: 95343
Conc: 1.10 ng/ml



#35 AR-1260-5


R.T.: 8.368 min
Delta R.T.: -0.006 min
Response: 257472
Conc: 3.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.289 min
Delta R.T.: -0.005 min
Response: 804018
Conc: 4.36 ng/ml




#36 AR-1262-1

R.T.: 7.834 min
Delta R.T.: 0.009 min
Response: 23156
Conc: 0.36 ng/ml



#36 AR-1262-1


R.T.: 6.844 min
Delta R.T.: 0.000 min
Response: 61354
Conc: 1.07 ng/ml



#37 AR-1262-2


R.T.: 8.368 min
Delta R.T.: -0.003 min
Response: 257472
Conc: 2.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.049 min
Delta R.T.: 0.000 min
Response: 95343
Conc: 0.81 ng/ml




#38 AR-1262-3

R.T.: 8.701 min
Delta R.T.: 0.012 min
Response: 179934
Conc: 2.80 ng/ml



#38 AR-1262-3


R.T.: 7.574 min
Delta R.T.: -0.004 min
Response: 11605
Conc: 0.14 ng/ml



#39 AR-1262-4


R.T.: 8.773 min
Delta R.T.: -0.005 min
Response: 20017
Conc: 0.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.635 min
Delta R.T.: -0.007 min
Response: 660240
Conc: 4.37 ng/ml




#40 AR-1262-5

R.T.: 9.431 min
Delta R.T.: -0.011 min
Response: 18346
Conc: 0.61 ng/ml



#40 AR-1262-5


R.T.: 8.138 min
Delta R.T.: -0.001 min
Response: 135035
Conc: 2.12 ng/ml



#41 AR-1268-1


R.T.: 8.701 min
Delta R.T.: 0.016 min
Response: 179934
Conc: 1.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.574 min
Delta R.T.: -0.004 min
Response: 11605
Conc: 0.05 ng/ml




#42 AR-1268-2

R.T.: 8.784 min
Delta R.T.: 0.003 min
Response: 25528
Conc: 0.26 ng/ml



#42 AR-1268-2


R.T.: 7.635 min
Delta R.T.: -0.007 min
Response: 660240
Conc: 3.01 ng/ml



#43 AR-1268-3


R.T.: 9.005 min
Delta R.T.: -0.007 min
Response: 52561
Conc: 0.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.851 min
Delta R.T.: 0.001 min
Response: 240212
Conc: 1.23 ng/ml




#44 AR-1268-4

R.T.: 9.431 min
Delta R.T.: -0.011 min
Response: 18346
Conc: 0.58 ng/ml



#44 AR-1268-4


R.T.: 8.138 min
Delta R.T.: 0.000 min
Response: 135035
Conc: 1.94 ng/ml



#45 AR-1268-5

R.T.: 9.864 min
Delta R.T.: 0.000 min
Response: 66591
Conc: 0.25 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.428 min
Delta R.T.: -0.003 min
Response: 205157
Conc: 0.40 ng/ml