

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080823\
 Data File : PP059222.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 00:15
 Operator : YP\AJ
 Sample : 03572-07
 Misc : AR1262 LOD 25 PPB
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2023

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/09/2023
 Supervised By :Ankita Jodhani 08/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 02:44:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 02:30:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.415	3.648	24019553	40897029	19.575	20.185
2) SA Decachlor...	10.210	8.685	17739538	33596544	23.243m	23.881
Target Compounds						
36) L8 AR-1262-1	7.825	6.843	1872951	1474799	29.407	25.633
37) L8 AR-1262-2	8.371	7.048	2347780	2834498	25.565	24.022
38) L8 AR-1262-3	8.688	7.577	1680863	2412627	26.119	28.351
39) L8 AR-1262-4	8.779	7.640	803487	3985299	24.196	26.382
40) L8 AR-1262-5	9.443	8.137	652150	1789060	21.854	28.052m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080823\
Data File : PP059222.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2023 00:15
Operator : YP\AJ
Sample : 03572-07
Misc : AR1262 LOD 25 PPB
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2023

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/09/2023
Supervised By :Ankita Jodhani 08/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 02:44:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 09 02:30:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

