

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080921\
 Data File : PP038157.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Aug 2021 19:02
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 00:55:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 06 03:08:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.893	3.879	2037051	1250897	45.837	48.568
2)	SA Decachlor...	10.882	9.145	1054458	1198896	46.353	46.338
Target Compounds							
16)	L4 AR-1242-1	6.210	5.124	498388	334395	440.296	464.046
17)	L4 AR-1242-2	6.234	5.143	706113	461521	438.845	461.754
18)	L4 AR-1242-3	6.299	5.333	448466	253968	438.061	459.082
19)	L4 AR-1242-4	6.407	5.431	370388	241385	451.575	459.448
20)	L4 AR-1242-5	7.189	5.992	360313	304959	436.950	471.960



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080921\
Data File : PP038157.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Aug 2021 19:02
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 00:55:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 06 03:08:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.893 min
Delta R.T.: -0.001 min
Response: 2037051
Conc: 45.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.879 min
Delta R.T.: 0.000 min
Response: 1250897
Conc: 48.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.882 min
Delta R.T.: -0.004 min
Response: 1054458
Conc: 46.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.145 min
Delta R.T.: -0.004 min
Response: 1198896
Conc: 46.34 ng/ml



#16 AR-1242-1


R.T.: 6.210 min
Delta R.T.: -0.002 min
Response: 498388
Conc: 440.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 334395
Conc: 464.05 ng/ml




#17 AR-1242-2

R.T.: 6.234 min
Delta R.T.: -0.001 min
Response: 706113
Conc: 438.84 ng/ml



#17 AR-1242-2


R.T.: 5.143 min
Delta R.T.: -0.003 min
Response: 461521
Conc: 461.75 ng/ml



#18 AR-1242-3


R.T.: 6.299 min
Delta R.T.: -0.002 min
Response: 448466
Conc: 438.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.333 min
Delta R.T.: -0.003 min
Response: 253968
Conc: 459.08 ng/ml




#19 AR-1242-4

R.T.: 6.407 min
Delta R.T.: -0.002 min
Response: 370388
Conc: 451.57 ng/ml



#19 AR-1242-4


R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 241385
Conc: 459.45 ng/ml



#20 AR-1242-5

R.T.: 7.189 min
Delta R.T.: -0.002 min
Response: 360313
Conc: 436.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.992 min
Delta R.T.: -0.003 min
Response: 304959
Conc: 471.96 ng/ml