

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
 Data File : PP050425.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2022 16:08
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 20:44:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.592	80701020	159.7E6	50.135	50.971
2) SA Decachlor...	10.094	8.557	85373704	74845795	48.483	48.630
Target Compounds						
21) L5 AR-1248-1	5.523	4.664	18970686	30918601	469.836	462.615
22) L5 AR-1248-2	5.795	4.898	26855490	40196207	480.710	465.319
23) L5 AR-1248-3	6.001	4.940	30092254	41518952	470.415	457.988
24) L5 AR-1248-4	6.408	5.110	36899026	48505246	466.467	460.207
25) L5 AR-1248-5	6.447	5.501	35559762	45604521	468.421	459.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
Data File : PP050425.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2022 16:08
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 20:44:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 80701020
Conc: 50.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: 0.000 min
Response: 159651540
Conc: 50.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.005 min
Response: 85373704
Conc: 48.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.557 min
Delta R.T.: 0.002 min
Response: 74845795
Conc: 48.63 ng/ml



#21 AR-1248-1


R.T.: 5.523 min
Delta R.T.: 0.002 min
Response: 18970686
Conc: 469.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.664 min
Delta R.T.: 0.000 min
Response: 30918601
Conc: 462.61 ng/ml




#22 AR-1248-2

R.T.: 5.795 min
Delta R.T.: 0.001 min
Response: 26855490
Conc: 480.71 ng/ml



#22 AR-1248-2


R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 40196207
Conc: 465.32 ng/ml



#23 AR-1248-3


R.T.: 6.001 min
Delta R.T.: 0.001 min
Response: 30092254
Conc: 470.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.940 min
Delta R.T.: 0.000 min
Response: 41518952
Conc: 457.99 ng/ml




#24 AR-1248-4

R.T.: 6.408 min
Delta R.T.: 0.002 min
Response: 36899026
Conc: 466.47 ng/ml



#24 AR-1248-4


R.T.: 5.110 min
Delta R.T.: 0.000 min
Response: 48505246
Conc: 460.21 ng/ml



#25 AR-1248-5

R.T.: 6.447 min
Delta R.T.: 0.002 min
Response: 35559762
Conc: 468.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 45604521
Conc: 459.31 ng/ml