

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
 Data File : PP050437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 00:06
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 01:16:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.591	82137471	163.5E6	51.028	52.209
2) SA Decachlor...	10.089	8.552	81765440	73078198	46.434	47.482
Target Compounds						
21) L5 AR-1248-1	5.521	4.662	19666290	30923540	487.063	462.689
22) L5 AR-1248-2	5.793	4.896	27145075	40795655	485.894	472.258
23) L5 AR-1248-3	5.999	4.937	30635942	42363943	478.914	467.309
24) L5 AR-1248-4	6.406	5.108	36651601	49110230	463.339	465.947
25) L5 AR-1248-5	6.444	5.498	35237929	46239793	464.181	465.710



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
Data File : PP050437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 00:06
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 01:16:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 82137471
Conc: 51.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.002 min
Response: 163529291
Conc: 52.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.089 min
Delta R.T.: 0.000 min
Response: 81765440
Conc: 46.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.552 min
Delta R.T.: -0.003 min
Response: 73078198
Conc: 47.48 ng/ml



#21 AR-1248-1


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 19666290
Conc: 487.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.662 min
Delta R.T.: -0.002 min
Response: 30923540
Conc: 462.69 ng/ml




#22 AR-1248-2

R.T.: 5.793 min
Delta R.T.: -0.001 min
Response: 27145075
Conc: 485.89 ng/ml



#22 AR-1248-2


R.T.: 4.896 min
Delta R.T.: -0.003 min
Response: 40795655
Conc: 472.26 ng/ml



#23 AR-1248-3


R.T.: 5.999 min
Delta R.T.: 0.000 min
Response: 30635942
Conc: 478.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.937 min
Delta R.T.: -0.003 min
Response: 42363943
Conc: 467.31 ng/ml




#24 AR-1248-4

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 36651601
Conc: 463.34 ng/ml



#24 AR-1248-4


R.T.: 5.108 min
Delta R.T.: -0.003 min
Response: 49110230
Conc: 465.95 ng/ml



#25 AR-1248-5

R.T.: 6.444 min
Delta R.T.: 0.000 min
Response: 35237929
Conc: 464.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.498 min
Delta R.T.: -0.003 min
Response: 46239793
Conc: 465.71 ng/ml