

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
 Data File : PP050445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 02:50
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 03:55:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.344	3.591	80925503	156.6E6	50.275	49.986
2)	SA Decachlor...	10.091	8.553	78958632	70898375	44.840	46.065
Target Compounds							
16)	L4 AR-1242-1	5.520	4.662	24771942	39460111	488.773	475.998
17)	L4 AR-1242-2	5.543	4.681	34829618	55342054	482.141	473.901
18)	L4 AR-1242-3	5.605	4.856	20922136	28946063	473.101	472.045
19)	L4 AR-1242-4	5.706	4.938	18010980	27039857	483.332	461.907
20)	L4 AR-1242-5	6.444	5.458	19676768	32608244	466.259	469.117



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
Data File : PP050445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 02:50
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 03:55:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 80925503
Conc: 50.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.002 min
Response: 156568039
Conc: 49.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.091 min
Delta R.T.: 0.001 min
Response: 78958632
Conc: 44.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.553 min
Delta R.T.: -0.002 min
Response: 70898375
Conc: 46.07 ng/ml



#16 AR-1242-1


R.T.: 5.520 min
Delta R.T.: -0.002 min
Response: 24771942
Conc: 488.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.662 min
Delta R.T.: -0.003 min
Response: 39460111
Conc: 476.00 ng/ml




#17 AR-1242-2

R.T.: 5.543 min
Delta R.T.: -0.002 min
Response: 34829618
Conc: 482.14 ng/ml



#17 AR-1242-2


R.T.: 4.681 min
Delta R.T.: -0.003 min
Response: 55342054
Conc: 473.90 ng/ml



#18 AR-1242-3


R.T.: 5.605 min
Delta R.T.: -0.002 min
Response: 20922136
Conc: 473.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.856 min
Delta R.T.: -0.003 min
Response: 28946063
Conc: 472.04 ng/ml




#19 AR-1242-4

R.T.: 5.706 min
Delta R.T.: -0.001 min
Response: 18010980
Conc: 483.33 ng/ml



#19 AR-1242-4


R.T.: 4.938 min
Delta R.T.: -0.002 min
Response: 27039857
Conc: 461.91 ng/ml



#20 AR-1242-5

R.T.: 6.444 min
Delta R.T.: -0.001 min
Response: 19676768
Conc: 466.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.458 min
Delta R.T.: -0.002 min
Response: 32608244
Conc: 469.12 ng/ml