

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
 Data File : PP050446.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 03:07
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 03:58:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.591	81760051	163.6E6	50.793	52.219
2) SA Decachlor...	10.090	8.553	81827076	73343755	46.469	47.654
Target Compounds						
21) L5 AR-1248-1	5.521	4.662	19657433	31793115	486.844	475.700
22) L5 AR-1248-2	5.793	4.896	27295346	40729938	488.583	471.497
23) L5 AR-1248-3	6.000	4.937	30473786	42205643	476.379	465.562
24) L5 AR-1248-4	6.407	5.108	36550408	48750400	462.060	462.533
25) L5 AR-1248-5	6.446	5.498	34974007	46077277	460.705	464.074



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081022\
Data File : PP050446.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 03:07
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 03:58:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: -0.001 min
Response: 81760051
Conc: 50.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.002 min
Response: 163562656
Conc: 52.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.090 min
Delta R.T.: 0.000 min
Response: 81827076
Conc: 46.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.553 min
Delta R.T.: -0.002 min
Response: 73343755
Conc: 47.65 ng/ml



#21 AR-1248-1


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 19657433
Conc: 486.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.662 min
Delta R.T.: -0.002 min
Response: 31793115
Conc: 475.70 ng/ml




#22 AR-1248-2

R.T.: 5.793 min
Delta R.T.: 0.000 min
Response: 27295346
Conc: 488.58 ng/ml



#22 AR-1248-2


R.T.: 4.896 min
Delta R.T.: -0.003 min
Response: 40729938
Conc: 471.50 ng/ml



#23 AR-1248-3


R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 30473786
Conc: 476.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.937 min
Delta R.T.: -0.003 min
Response: 42205643
Conc: 465.56 ng/ml




#24 AR-1248-4

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 36550408
Conc: 462.06 ng/ml



#24 AR-1248-4


R.T.: 5.108 min
Delta R.T.: -0.002 min
Response: 48750400
Conc: 462.53 ng/ml



#25 AR-1248-5

R.T.: 6.446 min
Delta R.T.: 0.000 min
Response: 34974007
Conc: 460.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.498 min
Delta R.T.: -0.003 min
Response: 46077277
Conc: 464.07 ng/ml