

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081023\
 Data File : PP059302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2023 13:54
 Operator : YP\AJ
 Sample : 03572-03
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 21:49:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 02:30:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.416	3.647	20302603	34800718	16.546	17.176
2)	SA Decachlor...	10.206	8.682	15449644	29852722	20.242	21.220
Target Compounds							
3)	L1 AR-1016-1	5.591	4.740	66922	230637	1.703	3.596 #
4)	L1 AR-1016-2	5.618	4.760	163436	131524	2.883	1.448 #
5)	L1 AR-1016-3	5.674	4.936	127416	110544	3.552	2.301 #
6)	L1 AR-1016-4	5.776	4.979	31908	218688	1.102	5.296 #
7)	L1 AR-1016-5	6.072	5.192	109911	352443	3.632	6.774 #
8)	L2 AR-1221-1	4.626	3.864	112573	25037	7.146	0.922 #
9)	L2 AR-1221-2	4.726	3.949	557618	769414	47.220	38.645
10)	L2 AR-1221-3	4.789	4.025	349655	120285	9.930	2.020 #
11)	L3 AR-1232-1	4.789	4.025	349655	120285	11.894	2.535 #
12)	L3 AR-1232-2	5.324	4.760	61025	131524	3.915	3.101
13)	L3 AR-1232-3	5.618	4.936	163436	110544	6.189	4.982
14)	L3 AR-1232-4	5.776	5.020	31908	402216	2.411	19.049 #
15)	L3 AR-1232-5	5.865	5.192	58047	352443	5.085	15.375 #
16)	L4 AR-1242-1	5.591	4.740	66922	230637	1.955	4.121 #
17)	L4 AR-1242-2	5.618	4.760	163436	131524	3.331	1.649 #
18)	L4 AR-1242-3	5.674	4.936	127416	110544	4.055	2.594 #
19)	L4 AR-1242-4	5.776	5.020	31908	402216	1.244	9.211 #
20)	L4 AR-1242-5	6.517	5.546	165022	724277	6.388	14.581 #
21)	L5 AR-1248-1	5.591	4.740	66922	230637	2.637	5.606 #
22)	L5 AR-1248-2	5.865	4.979	58047	218688	1.500	3.632 #
23)	L5 AR-1248-3	6.072	5.020	109911	402216	2.602	6.528 #
24)	L5 AR-1248-4	6.473	5.192	135840	352443	3.263	4.895 #
25)	L5 AR-1248-5	6.517	5.601	165022	89199	4.046	1.352 #
26)	L6 AR-1254-1	6.451	5.546	460829	724277	9.865	7.149 #
27)	L6 AR-1254-2	6.670	5.695	386706	1011541	5.620	11.459 #
28)	L6 AR-1254-3	7.042	6.091	605544	73510	8.773	0.532 #
29)	L6 AR-1254-4	7.329	6.331	101938	98512	2.264	1.226 #
30)	L6 AR-1254-5	7.743	6.748	1084658	1529251	23.084	12.788 #
31)	L7 AR-1260-1	7.205	6.229	1307431	2277121	23.898	22.992
32)	L7 AR-1260-2	7.463	6.420	1057055	2630957	17.754	22.738 #
33)	L7 AR-1260-3	7.823	6.574	1596455	2563811	40.566	23.049 #
34)	L7 AR-1260-4	8.049	7.046	1174906	2464393	26.454	28.426
35)	L7 AR-1260-5	8.370	7.290	1980443	5437905	24.765	29.502
36)	L8 AR-1262-1	7.823	6.840	1596455	1453636	25.065	25.265
37)	L8 AR-1262-2	8.370	7.046	1980443	2464393	21.565	20.885
38)	L8 AR-1262-3	8.685	7.575	1338270	2376076	20.796	27.921 #
39)	L8 AR-1262-4	8.777	7.638	567334	3802695	17.085	25.173 #
40)	L8 AR-1262-5	9.441	8.136	661511	1771298	22.168	27.774 #
41)	L9 AR-1268-1	8.685	7.575	1338270	2376076	11.784	9.744

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081023\
 Data File : PP059302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2023 13:54
 Operator : YP\AJ
 Sample : 03572-03
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 21:49:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 02:30:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.777	7.638	567334	3802695	5.765	17.362 #
43) L9	AR-1268-3	9.007	7.849	122528	268632	1.322	1.374
44) L9	AR-1268-4	9.441	8.136	661511	1771298	20.737	25.453
45) L9	AR-1268-5	9.862	8.427	389318	658896	1.434	1.271



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081023\
Data File : PP059302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2023 13:54
Operator : YP\AJ
Sample : 03572-03
Misc : AR1262 MDL 25 PPB
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 21:49:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 09 02:30:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 20302603
Conc: 16.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.002 min
Response: 34800718
Conc: 17.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.005 min
Response: 15449644
Conc: 20.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.682 min
Delta R.T.: -0.006 min
Response: 29852722
Conc: 21.22 ng/ml



#3 AR-1016-1


R.T.: 5.591 min
Delta R.T.: -0.003 min
Response: 66922
Conc: 1.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.740 min
Delta R.T.: -0.003 min
Response: 230637
Conc: 3.60 ng/ml




#4 AR-1016-2

R.T.: 5.618 min
Delta R.T.: 0.002 min
Response: 163436
Conc: 2.88 ng/ml



#4 AR-1016-2


R.T.: 4.760 min
Delta R.T.: -0.001 min
Response: 131524
Conc: 1.45 ng/ml



#5 AR-1016-3


R.T.: 5.674 min
Delta R.T.: -0.005 min
Response: 127416
Conc: 3.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.936 min
Delta R.T.: -0.002 min
Response: 110544
Conc: 2.30 ng/ml




#6 AR-1016-4

R.T.: 5.776 min
Delta R.T.: -0.002 min
Response: 31908
Conc: 1.10 ng/ml



#6 AR-1016-4


R.T.: 4.979 min
Delta R.T.: -0.001 min
Response: 218688
Conc: 5.30 ng/ml



#7 AR-1016-5


R.T.: 6.072 min
Delta R.T.: -0.002 min
Response: 109911
Conc: 3.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.192 min
Delta R.T.: -0.002 min
Response: 352443
Conc: 6.77 ng/ml




#8 AR-1221-1

R.T.: 4.626 min
Delta R.T.: 0.000 min
Response: 112573
Conc: 7.15 ng/ml



#8 AR-1221-1


R.T.: 3.864 min
Delta R.T.: 0.000 min
Response: 25037
Conc: 0.92 ng/ml



#9 AR-1221-2


R.T.: 4.726 min
Delta R.T.: 0.013 min
Response: 557618
Conc: 47.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.949 min
Delta R.T.: -0.002 min
Response: 769414
Conc: 38.64 ng/ml




#10 AR-1221-3

R.T.: 4.789 min
Delta R.T.: -0.002 min
Response: 349655
Conc: 9.93 ng/ml



#10 AR-1221-3


R.T.: 4.025 min
Delta R.T.: -0.002 min
Response: 120285
Conc: 2.02 ng/ml



#11 AR-1232-1


R.T.: 4.789 min
Delta R.T.: -0.001 min
Response: 349655
Conc: 11.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.025 min
Delta R.T.: -0.002 min
Response: 120285
Conc: 2.54 ng/ml




#12 AR-1232-2

R.T.: 5.324 min
Delta R.T.: 0.000 min
Response: 61025
Conc: 3.92 ng/ml



#12 AR-1232-2


R.T.: 4.760 min
Delta R.T.: -0.001 min
Response: 131524
Conc: 3.10 ng/ml



#13 AR-1232-3


R.T.: 5.618 min
Delta R.T.: 0.002 min
Response: 163436
Conc: 6.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.936 min
Delta R.T.: -0.001 min
Response: 110544
Conc: 4.98 ng/ml




#14 AR-1232-4

R.T.: 5.776 min
Delta R.T.: -0.002 min
Response: 31908
Conc: 2.41 ng/ml



#14 AR-1232-4


R.T.: 5.020 min
Delta R.T.: -0.002 min
Response: 402216
Conc: 19.05 ng/ml



#15 AR-1232-5


R.T.: 5.865 min
Delta R.T.: -0.004 min
Response: 58047
Conc: 5.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.192 min
Delta R.T.: -0.001 min
Response: 352443
Conc: 15.38 ng/ml




#16 AR-1242-1

R.T.: 5.591 min
Delta R.T.: -0.003 min
Response: 66922
Conc: 1.95 ng/ml



#16 AR-1242-1


R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 230637
Conc: 4.12 ng/ml



#17 AR-1242-2


R.T.: 5.618 min
Delta R.T.: 0.002 min
Response: 163436
Conc: 3.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.760 min
Delta R.T.: -0.001 min
Response: 131524
Conc: 1.65 ng/ml




#18 AR-1242-3

R.T.: 5.674 min
Delta R.T.: -0.005 min
Response: 127416
Conc: 4.06 ng/ml



#18 AR-1242-3


R.T.: 4.936 min
Delta R.T.: 0.000 min
Response: 110544
Conc: 2.59 ng/ml



#19 AR-1242-4


R.T.: 5.776 min
Delta R.T.: -0.001 min
Response: 31908
Conc: 1.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.020 min
Delta R.T.: -0.001 min
Response: 402216
Conc: 9.21 ng/ml




#20 AR-1242-5

R.T.: 6.517 min
Delta R.T.: -0.002 min
Response: 165022
Conc: 6.39 ng/ml



#20 AR-1242-5


R.T.: 5.546 min
Delta R.T.: -0.002 min
Response: 724277
Conc: 14.58 ng/ml



#21 AR-1248-1


R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 66922
Conc: 2.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 230637
Conc: 5.61 ng/ml




#22 AR-1248-2

R.T.: 5.865 min
Delta R.T.: -0.004 min
Response: 58047
Conc: 1.50 ng/ml



#22 AR-1248-2


R.T.: 4.979 min
Delta R.T.: -0.001 min
Response: 218688
Conc: 3.63 ng/ml



#23 AR-1248-3


R.T.: 6.072 min
Delta R.T.: -0.002 min
Response: 109911
Conc: 2.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.020 min
Delta R.T.: -0.001 min
Response: 402216
Conc: 6.53 ng/ml




#24 AR-1248-4

R.T.: 6.473 min
Delta R.T.: -0.007 min
Response: 135840
Conc: 3.26 ng/ml



#24 AR-1248-4


R.T.: 5.192 min
Delta R.T.: 0.000 min
Response: 352443
Conc: 4.90 ng/ml



#25 AR-1248-5


R.T.: 6.517 min
Delta R.T.: -0.002 min
Response: 165022
Conc: 4.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.601 min
Delta R.T.: 0.011 min
Response: 89199
Conc: 1.35 ng/ml




#26 AR-1254-1

R.T.: 6.451 min
Delta R.T.: -0.002 min
Response: 460829
Conc: 9.86 ng/ml



#26 AR-1254-1


R.T.: 5.546 min
Delta R.T.: -0.002 min
Response: 724277
Conc: 7.15 ng/ml



#27 AR-1254-2


R.T.: 6.670 min
Delta R.T.: -0.003 min
Response: 386706
Conc: 5.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.695 min
Delta R.T.: -0.002 min
Response: 1011541
Conc: 11.46 ng/ml




#28 AR-1254-3

R.T.: 7.042 min
Delta R.T.: 0.001 min
Response: 605544
Conc: 8.77 ng/ml



#28 AR-1254-3


R.T.: 6.091 min
Delta R.T.: -0.011 min
Response: 73510
Conc: 0.53 ng/ml



#29 AR-1254-4


R.T.: 7.329 min
Delta R.T.: 0.000 min
Response: 101938
Conc: 2.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.331 min
Delta R.T.: -0.002 min
Response: 98512
Conc: 1.23 ng/ml




#30 AR-1254-5

R.T.: 7.743 min
Delta R.T.: -0.005 min
Response: 1084658
Conc: 23.08 ng/ml



#30 AR-1254-5


R.T.: 6.748 min
Delta R.T.: -0.004 min
Response: 1529251
Conc: 12.79 ng/ml



#31 AR-1260-1


R.T.: 7.205 min
Delta R.T.: -0.002 min
Response: 1307431
Conc: 23.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.229 min
Delta R.T.: -0.004 min
Response: 2277121
Conc: 22.99 ng/ml




#32 AR-1260-2

R.T.: 7.463 min
Delta R.T.: -0.001 min
Response: 1057055
Conc: 17.75 ng/ml



#32 AR-1260-2


R.T.: 6.420 min
Delta R.T.: -0.004 min
Response: 2630957
Conc: 22.74 ng/ml



#33 AR-1260-3


R.T.: 7.823 min
Delta R.T.: -0.003 min
Response: 1596455
Conc: 40.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.574 min
Delta R.T.: -0.002 min
Response: 2563811
Conc: 23.05 ng/ml




#34 AR-1260-4

R.T.: 8.049 min
Delta R.T.: -0.002 min
Response: 1174906
Conc: 26.45 ng/ml



#34 AR-1260-4


R.T.: 7.046 min
Delta R.T.: -0.004 min
Response: 2464393
Conc: 28.43 ng/ml



#35 AR-1260-5


R.T.: 8.370 min
Delta R.T.: -0.004 min
Response: 1980443
Conc: 24.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.290 min
Delta R.T.: -0.004 min
Response: 5437905
Conc: 29.50 ng/ml




#36 AR-1262-1

R.T.: 7.823 min
Delta R.T.: -0.002 min
Response: 1596455
Conc: 25.07 ng/ml



#36 AR-1262-1


R.T.: 6.840 min
Delta R.T.: -0.003 min
Response: 1453636
Conc: 25.27 ng/ml



#37 AR-1262-2


R.T.: 8.370 min
Delta R.T.: -0.002 min
Response: 1980443
Conc: 21.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.046 min
Delta R.T.: -0.004 min
Response: 2464393
Conc: 20.89 ng/ml




#38 AR-1262-3

R.T.: 8.685 min
Delta R.T.: -0.004 min
Response: 1338270
Conc: 20.80 ng/ml



#38 AR-1262-3


R.T.: 7.575 min
Delta R.T.: -0.002 min
Response: 2376076
Conc: 27.92 ng/ml



#39 AR-1262-4


R.T.: 8.777 min
Delta R.T.: -0.001 min
Response: 567334
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.638 min
Delta R.T.: -0.004 min
Response: 3802695
Conc: 25.17 ng/ml




#40 AR-1262-5

R.T.: 9.441 min
Delta R.T.: 0.000 min
Response: 661511
Conc: 22.17 ng/ml



#40 AR-1262-5


R.T.: 8.136 min
Delta R.T.: -0.004 min
Response: 1771298
Conc: 27.77 ng/ml



#41 AR-1268-1


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 1338270
Conc: 11.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.575 min
Delta R.T.: -0.002 min
Response: 2376076
Conc: 9.74 ng/ml




#42 AR-1268-2

R.T.: 8.777 min
Delta R.T.: -0.004 min
Response: 567334
Conc: 5.76 ng/ml



#42 AR-1268-2


R.T.: 7.638 min
Delta R.T.: -0.004 min
Response: 3802695
Conc: 17.36 ng/ml



#43 AR-1268-3


R.T.: 9.007 min
Delta R.T.: -0.004 min
Response: 122528
Conc: 1.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.849 min
Delta R.T.: 0.000 min
Response: 268632
Conc: 1.37 ng/ml




#44 AR-1268-4

R.T.: 9.441 min
Delta R.T.: 0.000 min
Response: 661511
Conc: 20.74 ng/ml



#44 AR-1268-4


R.T.: 8.136 min
Delta R.T.: -0.003 min
Response: 1771298
Conc: 25.45 ng/ml



#45 AR-1268-5

R.T.: 9.862 min
Delta R.T.: -0.002 min
Response: 389318
Conc: 1.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.427 min
Delta R.T.: -0.004 min
Response: 658896
Conc: 1.27 ng/ml