Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081121\
Data File : PP038249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Aug 2021 15:46
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId : AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 01:05:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 06 03:08:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.892	3.878	2022453	1225773	45.509	47.593	
2)	SA Decachlor...	10.881	9.144	1064840	1165936	46.809	45.064	
	Target Compounds							
3)	L1 AR-1016-1	6.209	5.123	660704	442891	444.891	458.529	
4)	L1 AR-1016-2	6.233	5.142	943393	623021	442.966	465.047	
5)	L1 AR-1016-3	6.298	5.332	594110	342513	447.069	465.544	
6)	L1 AR-1016-4	6.406	5.386	494537	262700	455.491	462.466	
7)	L1 AR-1016-5	6.720	5.612	457405	334142	456.115	435.461	
8)	L2 AR-1221-1	5.134	4.131	70691	44823	130.418	149.007	
9)	L2 AR-1221-2	5.240	4.230	96395	64412	238.532	273.675	
10)	L2 AR-1221-3	5.327	4.318	401175	249084	333.024	337.778	
11)	L3 AR-1232-1	5.327	4.318	401175	249084	362.726	367.429	
12)	L3 AR-1232-2	5.914	5.142	509366	623021	933.580	1023.226	
13)	L3 AR-1232-3	6.233	5.332	943393	342513	958.510	1054.665	
14)	L3 AR-1232-4	6.406	5.430	494537	331826	989.720	1123.214	
15)	L3 AR-1232-5	6.504	5.612	370986	334142	1167.786	901.623	
16)	L4 AR-1242-1	6.209	5.123	660704	442891	583.692	614.608	
17)	L4 AR-1242-2	6.233	5.142	943393	623021	586.312	623.335	
18)	L4 AR-1242-3	6.298	5.332	594110	342513	580.325	619.139	
19)	L4 AR-1242-4	6.406	5.430	494537	331826	602.937	631.592	
20)	L4 AR-1242-5	7.188	5.991	86917	297591	105.403	460.556	#
21)	L5 AR-1248-1	6.209	5.123	660704	442891	777.599	836.871	
22)	L5 AR-1248-2	6.504	5.386	370986	262700	343.161	353.712	
23)	L5 AR-1248-3	6.720	5.430	457405	331826	358.926	421.715	
24)	L5 AR-1248-4	7.148	5.612	105297	334142	77.990	362.586	#
25)	L5 AR-1248-5	7.188	6.033	86917	51182	65.742	54.366	
26)	L6 AR-1254-1	7.119	5.991	379602	297591	291.328	222.438	
27)	L6 AR-1254-2	7.347	6.152	308842	227461	163.290	198.465	
28)	L6 AR-1254-3	7.730	6.569	252672	145383	127.358	77.008	#
29)	L6 AR-1254-4	8.025	6.809	114230	82039	79.596	68.139	
30)	L6 AR-1254-5	8.454	7.237	857252	809271	589.886	490.408	
31)	L7 AR-1260-1	7.896	6.706	715242	642868	498.363	493.753	
32)	L7 AR-1260-2	8.163	6.905	772419	717778	448.650	461.319	
33)	L7 AR-1260-3	8.527	7.058	544966	669484	456.182	439.811	
34)	L7 AR-1260-4	8.758	7.539	629744	569620	442.142	465.607	
35)	L7 AR-1260-5	9.083	7.789	1197158	1350279	450.895	462.358	
36)	L8 AR-1262-1	8.527	7.237	544966	809271	331.749	885.399	#
37)	L8 AR-1262-2	9.083	7.789	1197158	1350279	421.233	436.327	
38)	L8 AR-1262-3	9.390	8.075	241883	314541	204.237	248.633	
39)	L8 AR-1262-4	9.464f	8.136	475311	1027852	621.206	434.072	#
40)	L8 AR-1262-5	10.142	8.636	351185	383945	354.164	342.123	
41)	L9 AR-1268-1	9.390	8.075	241883	314541	74.033	86.678	
42)	L9 AR-1268-2	9.464f	8.136	475311	1027852	155.391	300.856	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081121\
Data File : PP038249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Aug 2021 15:46
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId : AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 01:05:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 06 03:08:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
43)	L9 AR-1268-3	0.000	8.345	0	15808	N.D.	5.548	#
44)	L9 AR-1268-4	10.142	8.636	351185	383945	318.253	303.374	
45)	L9 AR-1268-5	10.550	8.909	94517	105683	11.262	11.136	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081121\
Data File : PP038249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Aug 2021 15:46
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 01:05:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 06 03:08:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.892 min
Delta R.T.: -0.002 min
Response: 2022453
Conc: 45.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#1 Tetrachloro-m-xylene

R.T.: 3.878 min
Delta R.T.: 0.000 min
Response: 1225773
Conc: 47.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.881 min
Delta R.T.: -0.005 min
Response: 1064840
Conc: 46.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.144 min
Delta R.T.: -0.005 min
Response: 1165936
Conc: 45.06 ng/ml



#3 AR-1016-1

R.T.: 6.209 min
Delta R.T.: -0.003 min
Response: 660704
Conc: 444.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#3 AR-1016-1

R.T.: 5.123 min
Delta R.T.: -0.003 min
Response: 442891
Conc: 458.53 ng/ml



#4 AR-1016-2

R.T.: 6.233 min
Delta R.T.: -0.003 min
Response: 943393
Conc: 442.97 ng/ml



#4 AR-1016-2

R.T.: 5.142 min
Delta R.T.: -0.003 min
Response: 623021
Conc: 465.05 ng/ml



#5 AR-1016-3

R.T.: 6.298 min
Delta R.T.: -0.003 min
Response: 594110
Conc: 447.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#5 AR-1016-3

R.T.: 5.332 min
Delta R.T.: -0.003 min
Response: 342513
Conc: 465.54 ng/ml



#6 AR-1016-4

R.T.: 6.406 min
Delta R.T.: -0.003 min
Response: 494537
Conc: 455.49 ng/ml



#6 AR-1016-4

R.T.: 5.386 min
Delta R.T.: -0.003 min
Response: 262700
Conc: 462.47 ng/ml



#7 AR-1016-5

R.T.: 6.720 min
Delta R.T.: -0.003 min
Response: 457405
Conc: 456.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#7 AR-1016-5

R.T.: 5.612 min
Delta R.T.: -0.003 min
Response: 334142
Conc: 435.46 ng/ml



#8 AR-1221-1

R.T.: 5.134 min
Delta R.T.: -0.009 min
Response: 70691
Conc: 130.42 ng/ml



#8 AR-1221-1

R.T.: 4.131 min
Delta R.T.: -0.003 min
Response: 44823
Conc: 149.01 ng/ml



#9 AR-1221-2

R.T.: 5.240 min
Delta R.T.: -0.002 min
Response: 96395
Conc: 238.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#9 AR-1221-2

R.T.: 4.230 min
Delta R.T.: -0.002 min
Response: 64412
Conc: 273.67 ng/ml



#10 AR-1221-3

R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 401175
Conc: 333.02 ng/ml



#10 AR-1221-3

R.T.: 4.318 min
Delta R.T.: -0.002 min
Response: 249084
Conc: 337.78 ng/ml



#11 AR-1232-1

R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 401175
Conc: 362.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#11 AR-1232-1

R.T.: 4.318 min
Delta R.T.: -0.002 min
Response: 249084
Conc: 367.43 ng/ml



#12 AR-1232-2

R.T.: 5.914 min
Delta R.T.: -0.003 min
Response: 509366
Conc: 933.58 ng/ml



#12 AR-1232-2

R.T.: 5.142 min
Delta R.T.: -0.003 min
Response: 623021
Conc: 1023.23 ng/ml



#13 AR-1232-3

R.T.: 6.233 min
Delta R.T.: -0.002 min
Response: 943393
Conc: 958.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#13 AR-1232-3

R.T.: 5.332 min
Delta R.T.: -0.003 min
Response: 342513
Conc: 1054.66 ng/ml



#14 AR-1232-4

R.T.: 6.406 min
Delta R.T.: -0.003 min
Response: 494537
Conc: 989.72 ng/ml



#14 AR-1232-4

R.T.: 5.430 min
Delta R.T.: -0.003 min
Response: 331826
Conc: 1123.21 ng/ml



#15 AR-1232-5

R.T.: 6.504 min
Delta R.T.: -0.002 min
Response: 370986
Conc: 1167.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#15 AR-1232-5

R.T.: 5.612 min
Delta R.T.: -0.003 min
Response: 334142
Conc: 901.62 ng/ml



#16 AR-1242-1

R.T.: 6.209 min
Delta R.T.: -0.003 min
Response: 660704
Conc: 583.69 ng/ml



#16 AR-1242-1

R.T.: 5.123 min
Delta R.T.: -0.003 min
Response: 442891
Conc: 614.61 ng/ml



#17 AR-1242-2

R.T.: 6.233 min
Delta R.T.: -0.003 min
Response: 943393
Conc: 586.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#17 AR-1242-2

R.T.: 5.142 min
Delta R.T.: -0.003 min
Response: 623021
Conc: 623.34 ng/ml



#18 AR-1242-3

R.T.: 6.298 min
Delta R.T.: -0.003 min
Response: 594110
Conc: 580.32 ng/ml



#18 AR-1242-3

R.T.: 5.332 min
Delta R.T.: -0.003 min
Response: 342513
Conc: 619.14 ng/ml



#19 AR-1242-4

R.T.: 6.406 min
Delta R.T.: -0.003 min
Response: 494537
Conc: 602.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#19 AR-1242-4

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 331826
Conc: 631.59 ng/ml



#20 AR-1242-5

R.T.: 7.188 min
Delta R.T.: -0.003 min
Response: 86917
Conc: 105.40 ng/ml



#20 AR-1242-5

R.T.: 5.991 min
Delta R.T.: -0.004 min
Response: 297591
Conc: 460.56 ng/ml



#21 AR-1248-1

R.T.: 6.209 min
Delta R.T.: -0.003 min
Response: 660704
Conc: 777.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#21 AR-1248-1

R.T.: 5.123 min
Delta R.T.: -0.003 min
Response: 442891
Conc: 836.87 ng/ml



#22 AR-1248-2

R.T.: 6.504 min
Delta R.T.: -0.003 min
Response: 370986
Conc: 343.16 ng/ml



#22 AR-1248-2

R.T.: 5.386 min
Delta R.T.: -0.003 min
Response: 262700
Conc: 353.71 ng/ml



#23 AR-1248-3

R.T.: 6.720 min
Delta R.T.: -0.004 min
Response: 457405
Conc: 358.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#23 AR-1248-3

R.T.: 5.430 min
Delta R.T.: -0.003 min
Response: 331826
Conc: 421.72 ng/ml



#24 AR-1248-4

R.T.: 7.148 min
Delta R.T.: -0.003 min
Response: 105297
Conc: 77.99 ng/ml



#24 AR-1248-4

R.T.: 5.612 min
Delta R.T.: -0.003 min
Response: 334142
Conc: 362.59 ng/ml



#25 AR-1248-5

R.T.: 7.188 min
Delta R.T.: -0.004 min
Response: 86917
Conc: 65.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#25 AR-1248-5

R.T.: 6.033 min
Delta R.T.: -0.005 min
Response: 51182
Conc: 54.37 ng/ml



#26 AR-1254-1

R.T.: 7.119 min
Delta R.T.: -0.003 min
Response: 379602
Conc: 291.33 ng/ml



#26 AR-1254-1

R.T.: 5.991 min
Delta R.T.: -0.005 min
Response: 297591
Conc: 222.44 ng/ml



#27 AR-1254-2

R.T.: 7.347 min
Delta R.T.: -0.004 min
Response: 308842
Conc: 163.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#27 AR-1254-2

R.T.: 6.152 min
Delta R.T.: -0.004 min
Response: 227461
Conc: 198.46 ng/ml



#28 AR-1254-3

R.T.: 7.730 min
Delta R.T.: -0.003 min
Response: 252672
Conc: 127.36 ng/ml



#28 AR-1254-3

R.T.: 6.569 min
Delta R.T.: -0.004 min
Response: 145383
Conc: 77.01 ng/ml



#29 AR-1254-4

R.T.: 8.025 min
Delta R.T.: -0.005 min
Response: 114230
Conc: 79.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#29 AR-1254-4

R.T.: 6.809 min
Delta R.T.: -0.005 min
Response: 82039
Conc: 68.14 ng/ml



#30 AR-1254-5

R.T.: 8.454 min
Delta R.T.: -0.005 min
Response: 857252
Conc: 589.89 ng/ml



#30 AR-1254-5

R.T.: 7.237 min
Delta R.T.: -0.005 min
Response: 809271
Conc: 490.41 ng/ml



#31 AR-1260-1

R.T.: 7.896 min
Delta R.T.: -0.003 min
Response: 715242
Conc: 498.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#31 AR-1260-1

R.T.: 6.706 min
Delta R.T.: -0.004 min
Response: 642868
Conc: 493.75 ng/ml



#32 AR-1260-2

R.T.: 8.163 min
Delta R.T.: -0.003 min
Response: 772419
Conc: 448.65 ng/ml



#32 AR-1260-2

R.T.: 6.905 min
Delta R.T.: -0.004 min
Response: 717778
Conc: 461.32 ng/ml



#33 AR-1260-3

R.T.: 8.527 min
Delta R.T.: -0.003 min
Response: 544966
Conc: 456.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#33 AR-1260-3

R.T.: 7.058 min
Delta R.T.: -0.004 min
Response: 669484
Conc: 439.81 ng/ml



#34 AR-1260-4

R.T.: 8.758 min
Delta R.T.: -0.003 min
Response: 629744
Conc: 442.14 ng/ml



#34 AR-1260-4

R.T.: 7.539 min
Delta R.T.: -0.005 min
Response: 569620
Conc: 465.61 ng/ml



#35 AR-1260-5

R.T.: 9.083 min
Delta R.T.: -0.002 min
Response: 1197158
Conc: 450.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#35 AR-1260-5

R.T.: 7.789 min
Delta R.T.: -0.004 min
Response: 1350279
Conc: 462.36 ng/ml



#36 AR-1262-1

R.T.: 8.527 min
Delta R.T.: -0.004 min
Response: 544966
Conc: 331.75 ng/ml



#36 AR-1262-1

R.T.: 7.237 min
Delta R.T.: -0.003 min
Response: 809271
Conc: 885.40 ng/ml



#37 AR-1262-2

R.T.: 9.083 min
Delta R.T.: -0.003 min
Response: 1197158
Conc: 421.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#37 AR-1262-2

R.T.: 7.789 min
Delta R.T.: -0.005 min
Response: 1350279
Conc: 436.33 ng/ml



#38 AR-1262-3

R.T.: 9.390 min
Delta R.T.: -0.003 min
Response: 241883
Conc: 204.24 ng/ml



#38 AR-1262-3

R.T.: 8.075 min
Delta R.T.: -0.005 min
Response: 314541
Conc: 248.63 ng/ml



#39 AR-1262-4

R.T.: 9.464 min
Delta R.T.: -0.023 min
Response: 475311
Conc: 621.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#39 AR-1262-4

R.T.: 8.136 min
Delta R.T.: -0.007 min
Response: 1027852
Conc: 434.07 ng/ml



#40 AR-1262-5

R.T.: 10.142 min
Delta R.T.: -0.005 min
Response: 351185
Conc: 354.16 ng/ml



#40 AR-1262-5

R.T.: 8.636 min
Delta R.T.: -0.006 min
Response: 383945
Conc: 342.12 ng/ml



#41 AR-1268-1

R.T.: 9.390 min
Delta R.T.: -0.004 min
Response: 241883
Conc: 74.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#41 AR-1268-1

R.T.: 8.075 min
Delta R.T.: -0.005 min
Response: 314541
Conc: 86.68 ng/ml



#42 AR-1268-2

R.T.: 9.464 min
Delta R.T.: -0.026 min
Response: 475311
Conc: 155.39 ng/ml



#42 AR-1268-2

R.T.: 8.136 min
Delta R.T.: -0.008 min
Response: 1027852
Conc: 300.86 ng/ml



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 9.713 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#43 AR-1268-3

R.T.: 8.345 min
Delta R.T.: -0.005 min
Response: 15808
Conc: 5.55 ng/ml



#44 AR-1268-4

R.T.: 10.142 min
Delta R.T.: -0.006 min
Response: 351185
Conc: 318.25 ng/ml



#44 AR-1268-4

R.T.: 8.636 min
Delta R.T.: -0.005 min
Response: 383945
Conc: 303.37 ng/ml



#45 AR-1268-5

R.T.: 10.550 min
Delta R.T.: -0.007 min
Response: 94517
Conc: 11.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 8.909 min
Delta R.T.: -0.006 min
Response: 105683
Conc: 11.14 ng/ml