

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081122\
 Data File : PP050482.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2022 19:34
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 23:18:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.591	83388444	166.3E6	51.805	53.099
2) SA Decachlor...	10.092	8.555	85934520	77742313	48.802	50.512
Target Compounds						
26) L6 AR-1254-1	6.378	5.459	34660181	69639024	468.475	489.997
27) L6 AR-1254-2	6.597	5.605	53749919	58598084	462.676	495.144
28) L6 AR-1254-3	6.967	6.007	57274170	84470189	462.207	490.902
29) L6 AR-1254-4	7.255	6.235	44638683	53746808	456.219	509.411
30) L6 AR-1254-5	7.674	6.650	45472189	65983325	453.251	495.542



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081122\
Data File : PP050482.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2022 19:34
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 23:18:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: -0.001 min
Response: 83388444
Conc: 51.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.002 min
Response: 166317771
Conc: 53.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.092 min
Delta R.T.: 0.003 min
Response: 85934520
Conc: 48.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 77742313
Conc: 50.51 ng/ml



#26 AR-1254-1


R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 34660181
Conc: 468.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.459 min
Delta R.T.: -0.002 min
Response: 69639024
Conc: 490.00 ng/ml




#27 AR-1254-2

R.T.: 6.597 min
Delta R.T.: 0.000 min
Response: 53749919
Conc: 462.68 ng/ml



#27 AR-1254-2


R.T.: 5.605 min
Delta R.T.: -0.002 min
Response: 58598084
Conc: 495.14 ng/ml



#28 AR-1254-3


R.T.: 6.967 min
Delta R.T.: 0.000 min
Response: 57274170
Conc: 462.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.007 min
Delta R.T.: -0.001 min
Response: 84470189
Conc: 490.90 ng/ml




#29 AR-1254-4

R.T.: 7.255 min
Delta R.T.: 0.000 min
Response: 44638683
Conc: 456.22 ng/ml



#29 AR-1254-4


R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 53746808
Conc: 509.41 ng/ml



#30 AR-1254-5

R.T.: 7.674 min
Delta R.T.: 0.000 min
Response: 45472189
Conc: 453.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.650 min
Delta R.T.: -0.001 min
Response: 65983325
Conc: 495.54 ng/ml