

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081219\
 Data File : PP030323.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Aug 2019 19:19
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 07:07:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080919.M
 Quant Title : 8080.M
 QLast Update : Fri Aug 09 15:41:29 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.108	11.229	946.4E6	761.5E6	772.850	820.608
Target Compounds							
1) T	Dalapon	2.607	2.409	457.1E6	327.4E6	611.320	641.871
2) T	3,5-DICHL...	9.525	9.356	1515.8E6	904.4E6	727.774	754.911
3) T	4-Nitroph...	10.677	10.384	856.7E6	363.4E6	707.591	748.492
5) T	DICAMBA	11.418	11.524	4021.4E6	2769.7E6	745.499	764.011
6) T	MCPD	11.774	11.743	232.3E6	160.8E6	76.043	78.286
7) T	MCPA	12.027	12.130	291.8E6	224.1E6	72.905	73.622
8) T	DICHLORPROP	12.610	12.639	875.9E6	636.8E6	726.638	746.268
9) T	2,4-D	12.966	13.121	817.1E6	623.1E6	711.195	737.004
10) T	Pentachlo...	13.328	13.680	13780.7E6	9141.3E6	719.279	727.686
11) T	2,4,5-TP ...	14.215	14.281	4896.5E6	3633.6E6	723.030	744.271
12) T	2,4,5-T	14.610	14.822	3725.7E6	2751.9E6	686.665	722.607
13) T	2,4-DB	15.356	15.503	618.8E6	464.1E6	669.935	701.299
14) T	DINOSEB	16.818	15.910	2354.5E6	1738.1E6	573.860	585.634
15) T	Picloram	16.595	17.223	5512.7E6	4045.4E6	1849.716	2279.915
16) T	DCPA	17.171	17.121	6281.7E6	4572.4E6	755.368	753.953

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081219\
Data File : PP030323.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Aug 2019 19:19
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 07:07:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080919.M
Quant Title : 8080.M
QLast Update : Fri Aug 09 15:41:29 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

