

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081321\
 Data File : PP038352.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Aug 2021 12:00
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 00:25:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 06 03:08:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.891	3.876	2117339	1275487	47.644	49.523
2)	SA Decachlor...	10.880	9.139	1276263	1299073	56.103	50.210
Target Compounds							
21)	L5 AR-1248-1	6.209	5.120	395160	256428	465.074	484.537
22)	L5 AR-1248-2	6.503	5.383	485429	353680	449.020	476.210
23)	L5 AR-1248-3	6.720	5.427	572201	367836	449.007	467.480
24)	L5 AR-1248-4	7.149	5.610	607888	442932	450.242	480.637
25)	L5 AR-1248-5	7.188	6.031	600330	452984	454.078	481.162



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081321\
Data File : PP038352.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Aug 2021 12:00
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 00:25:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080521.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 06 03:08:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.891 min
Delta R.T.: -0.003 min
Response: 2117339
Conc: 47.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.876 min
Delta R.T.: -0.003 min
Response: 1275487
Conc: 49.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.880 min
Delta R.T.: -0.006 min
Response: 1276263
Conc: 56.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.139 min
Delta R.T.: -0.010 min
Response: 1299073
Conc: 50.21 ng/ml



#21 AR-1248-1


R.T.: 6.209 min
Delta R.T.: -0.003 min
Response: 395160
Conc: 465.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.120 min
Delta R.T.: -0.006 min
Response: 256428
Conc: 484.54 ng/ml




#22 AR-1248-2

R.T.: 6.503 min
Delta R.T.: -0.004 min
Response: 485429
Conc: 449.02 ng/ml



#22 AR-1248-2


R.T.: 5.383 min
Delta R.T.: -0.006 min
Response: 353680
Conc: 476.21 ng/ml



#23 AR-1248-3


R.T.: 6.720 min
Delta R.T.: -0.004 min
Response: 572201
Conc: 449.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.427 min
Delta R.T.: -0.007 min
Response: 367836
Conc: 467.48 ng/ml




#24 AR-1248-4

R.T.: 7.149 min
Delta R.T.: -0.003 min
Response: 607888
Conc: 450.24 ng/ml



#24 AR-1248-4


R.T.: 5.610 min
Delta R.T.: -0.006 min
Response: 442932
Conc: 480.64 ng/ml



#25 AR-1248-5

R.T.: 7.188 min
Delta R.T.: -0.004 min
Response: 600330
Conc: 454.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.031 min
Delta R.T.: -0.007 min
Response: 452984
Conc: 481.16 ng/ml