

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081522\
 Data File : PP050586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 20:20
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 23:21:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.590	80575882	164.7E6	50.058	52.585
2) SA Decachlor...	10.090	8.552	82361964	77050841	46.773	50.063
Target Compounds						
26) L6 AR-1254-1	6.377	5.456	35133230	72244642	474.868	508.331
27) L6 AR-1254-2	6.595	5.602	53675227	61287065	462.033	517.866
28) L6 AR-1254-3	6.966	6.004	57166260	88948056	461.336	516.926
29) L6 AR-1254-4	7.253	6.231	45736342	58072185	467.437	550.407
30) L6 AR-1254-5	7.672	6.647	45362185	71408091	452.154	536.282



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081522\
Data File : PP050586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 20:20
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 23:21:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 80575882
Conc: 50.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: -0.002 min
Response: 164708766
Conc: 52.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.090 min
Delta R.T.: 0.000 min
Response: 82361964
Conc: 46.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.552 min
Delta R.T.: -0.004 min
Response: 77050841
Conc: 50.06 ng/ml



#26 AR-1254-1


R.T.: 6.377 min
Delta R.T.: -0.002 min
Response: 35133230
Conc: 474.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.456 min
Delta R.T.: -0.004 min
Response: 72244642
Conc: 508.33 ng/ml




#27 AR-1254-2

R.T.: 6.595 min
Delta R.T.: -0.002 min
Response: 53675227
Conc: 462.03 ng/ml



#27 AR-1254-2


R.T.: 5.602 min
Delta R.T.: -0.004 min
Response: 61287065
Conc: 517.87 ng/ml



#28 AR-1254-3


R.T.: 6.966 min
Delta R.T.: -0.002 min
Response: 57166260
Conc: 461.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.004 min
Delta R.T.: -0.004 min
Response: 88948056
Conc: 516.93 ng/ml




#29 AR-1254-4

R.T.: 7.253 min
Delta R.T.: -0.003 min
Response: 45736342
Conc: 467.44 ng/ml



#29 AR-1254-4


R.T.: 6.231 min
Delta R.T.: -0.004 min
Response: 58072185
Conc: 550.41 ng/ml



#30 AR-1254-5

R.T.: 7.672 min
Delta R.T.: -0.002 min
Response: 45362185
Conc: 452.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.647 min
Delta R.T.: -0.004 min
Response: 71408091
Conc: 536.28 ng/ml