

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 04:13:12 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038506.D 1000 =PP038504.D 750 =PP038505.D
 250 =PP038507.D 50 =PP038508.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.136	3.931	4.005	4.271	4.560	4.181	E4	5.95
2) SA	Decachlorobiphenyl	2.598	2.395	2.471	2.649	2.835	2.590	E4	6.57
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	1.425	1.295	1.352	1.537	1.823	1.487	E3	14.05
37) L8	AR-1262-2	2.601	2.400	2.481	2.723	3.108	2.662	E3	10.41
38) L8	AR-1262-3	1.176	1.076	1.136	1.242	1.327	1.191	E3	8.14
39) L8	AR-1262-4	7.937	7.636	7.664	8.127	8.919	8.056	E2	6.49
40) L8	AR-1262-5	1.069	1.005	1.029	1.129	1.097	1.066	E3	4.73
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP038506.D 1000 =PP038504.D 750 =PP038505.D
 250 =PP038507.D 50 =PP038508.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.557	2.479	2.501	2.591	2.657	2.557	E4	2.80
2) SA	Decachlorobiphenyl	2.554	2.345	2.423	2.699	2.932	2.591	E4	9.01
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 04:13:12 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038506.D 1000 =PP038504.D 750 =PP038505.D
 250 =PP038507.D 50 =PP038508.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	9.065	8.419	8.676	9.375	9.659	9.039 E2	5.57
37) L8	AR-1262-2	3.085	2.921	2.986	3.151	3.305	3.090 E3	4.84
38) L8	AR-1262-3	1.237	1.141	1.177	1.296	1.430	1.256 E3	9.03
39) L8	AR-1262-4	2.357	2.213	2.266	2.443	2.538	2.363 E3	5.56
40) L8	AR-1262-5	1.147	1.065	1.098	1.171	1.205	1.137 E3	4.92
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range