

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 04:29:44 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038511.D 1000 =PP038509.D 750 =PP038510.D
 250 =PP038512.D 50 =PP038513.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.278	4.075	4.098	4.331	4.634	4.283	E4	5.26
2) SA	Decachlorobiphenyl	4.644	4.317	4.455	4.763	5.137	4.663	E4	6.76
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	3.158	2.960	3.005	3.276	3.639	3.208	E3	8.47
42) L9	AR-1268-2	3.069	2.876	2.917	3.132	3.411	3.081	E3	6.89
43) L9	AR-1268-3	2.657	2.481	2.526	2.721	2.908	2.659	E3	6.38
44) L9	AR-1268-4	1.220	1.143	1.172	1.232	1.324	1.218	E3	5.69
45) L9	AR-1268-5	0.956	0.907	0.925	0.962	1.014	0.953	E4	4.28

Signal #2 Calibration Files

500 =PP038511.D 1000 =PP038509.D 750 =PP038510.D
 250 =PP038512.D 50 =PP038513.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.646	2.561	2.554	2.624	2.671	2.611	E4	1.98
2) SA	Decachlorobiphenyl	4.485	4.159	4.276	4.668	5.252	4.568	E4	9.40
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 04:29:44 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038511.D 1000 =PP038509.D 750 =PP038510.D
 250 =PP038512.D 50 =PP038513.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.631	3.411	3.478	3.716	4.022	3.652 E3	6.57
42) L9	AR-1268-2	3.441	3.249	3.314	3.492	3.752	3.450 E3	5.64
43) L9	AR-1268-3	2.904	2.728	2.785	2.954	3.185	2.911 E3	6.11
44) L9	AR-1268-4	1.278	1.190	1.225	1.307	1.324	1.265 E3	4.45
45) L9	AR-1268-5	0.966	0.923	0.937	0.972	1.061	0.972 E4	5.54

(#) = Out of Range