

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 03:56:43 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038501.D 1000 =PP038499.D 750 =PP038500.D
 250 =PP038502.D 50 =PP038503.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.265	3.986	4.143	4.284	4.688	4.273 E4	6.10
2) SA	Decachlorobiphenyl	2.627	2.423	2.551	2.688	2.840	2.626 E4	5.92
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.148	1.006	1.068	1.216	1.268	1.141 E3	9.32
27) L6	AR-1254-2	1.694	1.494	1.586	1.796	2.065	1.727 E3	12.77
28) L6	AR-1254-3	1.776	1.574	1.660	1.880	2.085	1.795 E3	11.10
29) L6	AR-1254-4	1.275	1.127	1.192	1.352	1.506	1.290 E3	11.43
30) L6	AR-1254-5	1.268	1.142	1.194	1.327	1.253	1.237 E3	5.73
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PP038501.D 1000 =PP038499.D 750 =PP038500.D
 250 =PP038502.D 50 =PP038503.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.643	2.515	2.604	2.596	2.719	2.615 E4	2.84
2) SA	Decachlorobiphenyl	2.610	2.371	2.488	2.719	2.980	2.633 E4	8.87
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP081721.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 18 03:56:43 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP038501.D 1000 =PP038499.D 750 =PP038500.D
 250 =PP038502.D 50 =PP038503.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.319	1.195	1.259	1.352	1.442	1.313 E3	7.14
27) L6	AR-1254-2	1.144	1.028	1.086	1.182	1.298	1.148 E3	8.93
28) L6	AR-1254-3	1.863	1.713	1.798	1.918	2.011	1.861 E3	6.11
29) L6	AR-1254-4	1.192	1.100	1.155	1.208	1.188	1.169 E3	3.67
30) L6	AR-1254-5	1.608	1.515	1.586	1.638	1.735	1.616 E3	4.98
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range