

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081721\
 Data File : PP038518.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Aug 2021 00:37
 Operator : AJ\MA
 Sample : AR1262ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 04:13:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:13:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.885	3.871	2069345	1278286	49.497	49.993
2) SA Decachlor...	10.871	9.130	1323219	1285866	51.098	49.632
Target Compounds						
36) L8 AR-1262-1	8.523	7.226	719583	457431	484.065	506.076
37) L8 AR-1262-2	9.078	7.779	1318974	1536852	495.406	497.431
38) L8 AR-1262-3	9.383	8.064	607084	620433	509.520	493.823
39) L8 AR-1262-4	9.476	8.127	403561	1187305	500.920	502.376
40) L8 AR-1262-5	10.135	8.625	552867	578511	518.674	508.802



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081721\
Data File : PP038518.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Aug 2021 00:37
Operator : AJ\MA
Sample : AR1262ICV500
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 04:13:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:13:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.885 min
Delta R.T.: 0.000 min
Response: 2069345
Conc: 49.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.871 min
Delta R.T.: 0.000 min
Response: 1278286
Conc: 49.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.871 min
Delta R.T.: 0.000 min
Response: 1323219
Conc: 51.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.130 min
Delta R.T.: -0.002 min
Response: 1285866
Conc: 49.63 ng/ml



#36 AR-1262-1


R.T.: 8.523 min
Delta R.T.: 0.001 min
Response: 719583
Conc: 484.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 7.226 min
Delta R.T.: 0.000 min
Response: 457431
Conc: 506.08 ng/ml




#37 AR-1262-2

R.T.: 9.078 min
Delta R.T.: 0.002 min
Response: 1318974
Conc: 495.41 ng/ml



#37 AR-1262-2


R.T.: 7.779 min
Delta R.T.: 0.000 min
Response: 1536852
Conc: 497.43 ng/ml



#38 AR-1262-3


R.T.: 9.383 min
Delta R.T.: 0.003 min
Response: 607084
Conc: 509.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 620433
Conc: 493.82 ng/ml




#39 AR-1262-4

R.T.: 9.476 min
Delta R.T.: 0.000 min
Response: 403561
Conc: 500.92 ng/ml



#39 AR-1262-4


R.T.: 8.127 min
Delta R.T.: 0.000 min
Response: 1187305
Conc: 502.38 ng/ml



#40 AR-1262-5

R.T.: 10.135 min
Delta R.T.: 0.003 min
Response: 552867
Conc: 518.67 ng/ml

Instrument :
ECD_P
ClientSampleId :



#40 AR-1262-5

R.T.: 8.625 min
Delta R.T.: 0.000 min
Response: 578511
Conc: 508.80 ng/ml