

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081721\
 Data File : PP038490.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Aug 2021 16:42
 Operator : AJ\MA
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 02:29:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 02:29:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.886	3.872	3073678	1895649	74.744	74.812
2) SA Decachlor...	10.874	9.133	1882192	1833642	75.514	75.005
Target Compounds						
16) L4 AR-1242-1	6.205	5.116	702448	488010	744.387	743.180
17) L4 AR-1242-2	6.229	5.136	1014702	702578	749.563	748.507
18) L4 AR-1242-3	6.295	5.325	635184	380032	747.273	751.036
19) L4 AR-1242-4	6.402	5.423	527231	353281	748.957	751.776
20) L4 AR-1242-5	7.185	5.984	491708	452497	747.322	749.900



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081721\
Data File : PP038490.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Aug 2021 16:42
Operator : AJ\MA
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 02:29:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 02:29:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.886 min
Delta R.T.: 0.000 min
Response: 3073678
Conc: 74.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.872 min
Delta R.T.: 0.000 min
Response: 1895649
Conc: 74.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.874 min
Delta R.T.: 0.000 min
Response: 1882192
Conc: 75.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.133 min
Delta R.T.: 0.000 min
Response: 1833642
Conc: 75.01 ng/ml



#16 AR-1242-1


R.T.: 6.205 min
Delta R.T.: 0.000 min
Response: 702448
Conc: 744.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 488010
Conc: 743.18 ng/ml




#17 AR-1242-2

R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 1014702
Conc: 749.56 ng/ml



#17 AR-1242-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 702578
Conc: 748.51 ng/ml



#18 AR-1242-3


R.T.: 6.295 min
Delta R.T.: 0.000 min
Response: 635184
Conc: 747.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 380032
Conc: 751.04 ng/ml




#19 AR-1242-4

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 527231
Conc: 748.96 ng/ml



#19 AR-1242-4


R.T.: 5.423 min
Delta R.T.: 0.000 min
Response: 353281
Conc: 751.78 ng/ml



#20 AR-1242-5

R.T.: 7.185 min
Delta R.T.: 0.000 min
Response: 491708
Conc: 747.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.984 min
Delta R.T.: 0.000 min
Response: 452497
Conc: 749.90 ng/ml