

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081722\
 Data File : PP050676.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2022 00:57
 Operator : YP\AJ
 Sample : N4228-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 OK-03-081622

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/18/2022
 Supervised By :Ankita Jodhani 08/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 03:54:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.342	3.589	26830743	57018989	16.669m	18.204
2) SA Decachlor...	10.092	8.552	29286523	25239179	16.632	16.399

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081722\
Data File : PP050676.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2022 00:57
Operator : YP\AJ
Sample : N4228-01
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
OK-03-081622


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/18/2022
Supervised By :Ankita Jodhani 08/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 03:54:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.342 min
Delta R.T.: -0.004 min
Response: 26830743
Conc: 16.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-03-081622


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/18/2022
Supervised By :Ankita Jodhani 08/18/2022




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: -0.004 min
Response: 57018989
Conc: 18.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.092 min
Delta R.T.: 0.002 min
Response: 29286523
Conc: 16.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.552 min
Delta R.T.: -0.003 min
Response: 25239179
Conc: 16.40 ng/ml