Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
Data File : PP038540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Aug 2021 15:07
Operator : AJ\MA
Sample : M3435-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 02:51:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.885	3.871	649813	413545	16.101	17.257	
2)	SA Decachlor...	10.871	9.132	472763	409379	19.215	16.808	
	Target Compounds							
3)	L1 AR-1016-1	6.201	5.134f	95553	42456	73.635	48.707	#
4)	L1 AR-1016-2	6.230	5.134	172591	42456	93.393	34.138	#
5)	L1 AR-1016-3	6.295	5.325	94196	67970	82.443	99.818	
6)	L1 AR-1016-4	0.000	5.378	0	552185	N.D.	1054.426	#
7)	L1 AR-1016-5	6.714	5.604	441753	370597	521.127	558.754	
8)	L2 AR-1221-1	5.136	0.000	274073	0	586.060	N.D.	#
10)	L2 AR-1221-3	0.000	4.331f	0	98852	N.D.	146.401	#
11)	L3 AR-1232-1	0.000	4.331f	0	98852	N.D.	159.150	#
12)	L3 AR-1232-2	5.908	5.134	101830	42456	218.680	74.385	#
13)	L3 AR-1232-3	6.230	5.325	172591	67970	202.154	232.523	
14)	L3 AR-1232-4	0.000	5.420	0	290681	N.D.	1174.858	#
15)	L3 AR-1232-5	6.497	5.604	792016	370597	2947.440	1370.712	#
16)	L4 AR-1242-1	6.201	5.134f	95553	42456	95.930	61.491	#
17)	L4 AR-1242-2	6.230	5.134	172591	42456	120.106	43.492	#
18)	L4 AR-1242-3	6.295	5.325	94196	67970	105.217	129.257	
19)	L4 AR-1242-4	0.000	5.420	0	290681	N.D.	608.117	#
20)	L4 AR-1242-5	7.185	5.982	479324	1558730	670.422	2517.946	#
21)	L5 AR-1248-1	6.201	5.134f	95553	42456	119.919	77.913	#
22)	L5 AR-1248-2	6.497	5.378	792016	552185	807.811	742.962	
23)	L5 AR-1248-3	6.714	5.420	441753	290681	382.695	382.797	
24)	L5 AR-1248-4	7.143	5.604	513778	370597	416.062	400.662	
25)	L5 AR-1248-5	7.185	6.024	479324	170151	392.029	188.451	#
26)	L6 AR-1254-1	7.112	5.982	1565166	1558730	1371.727	1186.893	
27)	L6 AR-1254-2	7.342	6.143	2464306	1537935	1427.049	1339.928	
28)	L6 AR-1254-3	7.725	6.560	2800685	3259361	1560.165	1751.636	
29)	L6 AR-1254-4	8.021	6.800	1402913	1081413	1087.265	925.370	
30)	L6 AR-1254-5	8.449	7.227	3084498	3407002	2493.912	2107.886	
31)	L7 AR-1260-1	7.891	6.698	1776072	1975806	1484.019	1688.452	
32)	L7 AR-1260-2	8.157	6.896	3106683	2015268	2117.439	1415.746	#
33)	L7 AR-1260-3	8.522	7.046	1742894	1838928	1816.465	1278.670	#
34)	L7 AR-1260-4	8.751	7.529	2180426	1369672	1917.463	1219.592	#
35)	L7 AR-1260-5	9.077	7.779	2962855	3288886	1320.349	1214.202	
36)	L8 AR-1262-1	8.522	7.227	1742894	3407002	1172.448	3769.317	#
37)	L8 AR-1262-2	9.077	7.779	2962855	3288886	1112.847	1064.509	
38)	L8 AR-1262-3	9.383	8.064	3035807	2005487	2547.923	1596.233	#
39)	L8 AR-1262-4	9.477	8.129	2557718	3304939	3174.772	1398.395	#
40)	L8 AR-1262-5	10.135	8.626	1780533	1611831	1670.414	1417.611	
41)	L9 AR-1268-1	9.383	8.064	3035807	2005487	946.448	549.207	#
42)	L9 AR-1268-2	9.477	8.129	2557718	3304939	830.192	958.016	

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
Data File : PP038540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Aug 2021 15:07
Operator : AJ\MA
Sample : M3435-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 02:51:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	9.700	8.335	445521	429938	167.580	147.678
44) L9	AR-1268-4	10.135	8.626	1780533	1611831	1461.544	1274.458
45) L9	AR-1268-5	10.543	8.899	2279060	1990105	239.208	204.774

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
Data File : PP038540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Aug 2021 15:07
Operator : AJ\MA
Sample : M3435-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 02:51:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.885 min
Delta R.T.: -0.003 min
Response: 649813
Conc: 16.10 ng/ml

Instrument :
ECD_P
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 3.871 min
Delta R.T.: -0.003 min
Response: 413545
Conc: 17.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.871 min
Delta R.T.: -0.006 min
Response: 472763
Conc: 19.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.132 min
Delta R.T.: -0.004 min
Response: 409379
Conc: 16.81 ng/ml



#3 AR-1016-1

R.T.: 6.201 min
Delta R.T.: -0.005 min
Response: 95553
Conc: 73.64 ng/ml

Instrument :
ECD_P
ClientSampleId :



#3 AR-1016-1

R.T.: 5.134 min
Delta R.T.: 0.016 min
Response: 42456
Conc: 48.71 ng/ml



#4 AR-1016-2

R.T.: 6.230 min
Delta R.T.: 0.000 min
Response: 172591
Conc: 93.39 ng/ml



#4 AR-1016-2

R.T.: 5.134 min
Delta R.T.: -0.003 min
Response: 42456
Conc: 34.14 ng/ml



#5 AR-1016-3

R.T.: 6.295 min
Delta R.T.: -0.001 min
Response: 94196
Conc: 82.44 ng/ml

Instrument :
ECD_P
ClientSampleId :



#5 AR-1016-3

R.T.: 5.325 min
Delta R.T.: -0.002 min
Response: 67970
Conc: 99.82 ng/ml



#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 6.403 min
Response: 0
Conc: N.D.



#6 AR-1016-4

R.T.: 5.378 min
Delta R.T.: -0.004 min
Response: 552185
Conc: 1054.43 ng/ml



#7 AR-1016-5

R.T.: 6.714 min
Delta R.T.: -0.003 min
Response: 441753
Conc: 521.13 ng/ml

Instrument :
ECD_P
ClientSampleId :



#7 AR-1016-5

R.T.: 5.604 min
Delta R.T.: -0.003 min
Response: 370597
Conc: 558.75 ng/ml



#8 AR-1221-1

R.T.: 5.136 min
Delta R.T.: -0.003 min
Response: 274073
Conc: 586.06 ng/ml



#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.129 min
Response: 0
Conc: N.D.



#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.325 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#10 AR-1221-3

R.T.: 4.331 min
Delta R.T.: 0.017 min
Response: 98852
Conc: 146.40 ng/ml



#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 5.323 min
Response: 0
Conc: N.D.



#11 AR-1232-1

R.T.: 4.331 min
Delta R.T.: 0.019 min
Response: 98852
Conc: 159.15 ng/ml



#12 AR-1232-2

R.T.: 5.908 min
Delta R.T.: -0.002 min
Response: 101830
Conc: 218.68 ng/ml

Instrument :
ECD_P
ClientSampleId :



#12 AR-1232-2

R.T.: 5.134 min
Delta R.T.: -0.002 min
Response: 42456
Conc: 74.39 ng/ml



#13 AR-1232-3

R.T.: 6.230 min
Delta R.T.: 0.000 min
Response: 172591
Conc: 202.15 ng/ml



#13 AR-1232-3

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 67970
Conc: 232.52 ng/ml



#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 6.402 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#14 AR-1232-4

R.T.: 5.420 min
Delta R.T.: -0.003 min
Response: 290681
Conc: 1174.86 ng/ml



#15 AR-1232-5

R.T.: 6.497 min
Delta R.T.: -0.003 min
Response: 792016
Conc: 2947.44 ng/ml



#15 AR-1232-5

R.T.: 5.604 min
Delta R.T.: -0.001 min
Response: 370597
Conc: 1370.71 ng/ml



#16 AR-1242-1

R.T.: 6.201 min
Delta R.T.: -0.004 min
Response: 95553
Conc: 95.93 ng/ml

Instrument :
ECD_P
ClientSampleId :



#16 AR-1242-1

R.T.: 5.134 min
Delta R.T.: 0.018 min
Response: 42456
Conc: 61.49 ng/ml



#17 AR-1242-2

R.T.: 6.230 min
Delta R.T.: 0.000 min
Response: 172591
Conc: 120.11 ng/ml



#17 AR-1242-2

R.T.: 5.134 min
Delta R.T.: -0.002 min
Response: 42456
Conc: 43.49 ng/ml



#18 AR-1242-3

R.T.: 6.295 min
Delta R.T.: 0.000 min
Response: 94196
Conc: 105.22 ng/ml

Instrument :
ECD_P
ClientSampleId :



#18 AR-1242-3

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 67970
Conc: 129.26 ng/ml



#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 6.403 min
Response: 0
Conc: N.D.



#19 AR-1242-4

R.T.: 5.420 min
Delta R.T.: -0.003 min
Response: 290681
Conc: 608.12 ng/ml



#20 AR-1242-5

R.T.: 7.185 min
Delta R.T.: 0.000 min
Response: 479324
Conc: 670.42 ng/ml

Instrument :
ECD_P
ClientSampleId :



#20 AR-1242-5

R.T.: 5.982 min
Delta R.T.: -0.002 min
Response: 1558730
Conc: 2517.95 ng/ml



#21 AR-1248-1

R.T.: 6.201 min
Delta R.T.: -0.003 min
Response: 95553
Conc: 119.92 ng/ml



#21 AR-1248-1

R.T.: 5.134 min
Delta R.T.: 0.019 min
Response: 42456
Conc: 77.91 ng/ml



#22 AR-1248-2

R.T.: 6.497 min
Delta R.T.: -0.001 min
Response: 792016
Conc: 807.81 ng/ml

Instrument :
ECD_P
ClientSampleId :



#22 AR-1248-2

R.T.: 5.378 min
Delta R.T.: 0.000 min
Response: 552185
Conc: 742.96 ng/ml



#23 AR-1248-3

R.T.: 6.714 min
Delta R.T.: 0.000 min
Response: 441753
Conc: 382.70 ng/ml



#23 AR-1248-3

R.T.: 5.420 min
Delta R.T.: -0.002 min
Response: 290681
Conc: 382.80 ng/ml



#24 AR-1248-4

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 513778
Conc: 416.06 ng/ml

Instrument :
ECD_P
ClientSampleId :



#24 AR-1248-4

R.T.: 5.604 min
Delta R.T.: 0.000 min
Response: 370597
Conc: 400.66 ng/ml



#25 AR-1248-5

R.T.: 7.185 min
Delta R.T.: 0.002 min
Response: 479324
Conc: 392.03 ng/ml



#25 AR-1248-5

R.T.: 6.024 min
Delta R.T.: 0.000 min
Response: 170151
Conc: 188.45 ng/ml



#26 AR-1254-1

R.T.: 7.112 min
Delta R.T.: 0.000 min
Response: 1565166
Conc: 1371.73 ng/ml

Instrument :
ECD_P
ClientSampleId :



#26 AR-1254-1

R.T.: 5.982 min
Delta R.T.: 0.000 min
Response: 1558730
Conc: 1186.89 ng/ml



#27 AR-1254-2

R.T.: 7.342 min
Delta R.T.: -0.001 min
Response: 2464306
Conc: 1427.05 ng/ml



#27 AR-1254-2

R.T.: 6.143 min
Delta R.T.: 0.000 min
Response: 1537935
Conc: 1339.93 ng/ml



#28 AR-1254-3

R.T.: 7.725 min
Delta R.T.: 0.000 min
Response: 2800685
Conc: 1560.17 ng/ml

Instrument :
ECD_P
ClientSampleId :



#28 AR-1254-3

R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 3259361
Conc: 1751.64 ng/ml



#29 AR-1254-4

R.T.: 8.021 min
Delta R.T.: 0.000 min
Response: 1402913
Conc: 1087.26 ng/ml



#29 AR-1254-4

R.T.: 6.800 min
Delta R.T.: 0.000 min
Response: 1081413
Conc: 925.37 ng/ml



#30 AR-1254-5

R.T.: 8.449 min
Delta R.T.: -0.001 min
Response: 3084498
Conc: 2493.91 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 7.227 min
Delta R.T.: 0.000 min
Response: 3407002
Conc: 2107.89 ng/ml



#31 AR-1260-1

R.T.: 7.891 min
Delta R.T.: -0.003 min
Response: 1776072
Conc: 1484.02 ng/ml



#31 AR-1260-1

R.T.: 6.698 min
Delta R.T.: -0.003 min
Response: 1975806
Conc: 1688.45 ng/ml



#32 AR-1260-2

R.T.: 8.157 min
Delta R.T.: -0.002 min
Response: 3106683
Conc: 2117.44 ng/ml

Instrument :
ECD_P
ClientSampleId :



#32 AR-1260-2

R.T.: 6.896 min
Delta R.T.: -0.003 min
Response: 2015268
Conc: 1415.75 ng/ml



#33 AR-1260-3

R.T.: 8.522 min
Delta R.T.: -0.003 min
Response: 1742894
Conc: 1816.47 ng/ml



#33 AR-1260-3

R.T.: 7.046 min
Delta R.T.: -0.005 min
Response: 1838928
Conc: 1278.67 ng/ml



#34 AR-1260-4

R.T.: 8.751 min
Delta R.T.: -0.005 min
Response: 2180426
Conc: 1917.46 ng/ml

Instrument :
ECD_P
ClientSampleId :



#34 AR-1260-4

R.T.: 7.529 min
Delta R.T.: -0.003 min
Response: 1369672
Conc: 1219.59 ng/ml



#35 AR-1260-5

R.T.: 9.077 min
Delta R.T.: -0.004 min
Response: 2962855
Conc: 1320.35 ng/ml



#35 AR-1260-5

R.T.: 7.779 min
Delta R.T.: -0.004 min
Response: 3288886
Conc: 1214.20 ng/ml



#36 AR-1262-1

R.T.: 8.522 min
Delta R.T.: 0.000 min
Response: 1742894
Conc: 1172.45 ng/ml

Instrument :
ECD_P
ClientSampleId :



#36 AR-1262-1

R.T.: 7.227 min
Delta R.T.: 0.000 min
Response: 3407002
Conc: 3769.32 ng/ml



#37 AR-1262-2

R.T.: 9.077 min
Delta R.T.: 0.000 min
Response: 2962855
Conc: 1112.85 ng/ml



#37 AR-1262-2

R.T.: 7.779 min
Delta R.T.: 0.000 min
Response: 3288886
Conc: 1064.51 ng/ml



#38 AR-1262-3

R.T.: 9.383 min
Delta R.T.: 0.000 min
Response: 3035807
Conc: 2547.92 ng/ml

Instrument :
ECD_P
ClientSampleId :



#38 AR-1262-3

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 2005487
Conc: 1596.23 ng/ml



#39 AR-1262-4

R.T.: 9.477 min
Delta R.T.: 0.002 min
Response: 2557718
Conc: 3174.77 ng/ml



#39 AR-1262-4

R.T.: 8.129 min
Delta R.T.: 0.000 min
Response: 3304939
Conc: 1398.39 ng/ml



#40 AR-1262-5

R.T.: 10.135 min
Delta R.T.: 0.000 min
Response: 1780533
Conc: 1670.41 ng/ml

Instrument :
ECD_P
ClientSampleId :



#40 AR-1262-5

R.T.: 8.626 min
Delta R.T.: 0.000 min
Response: 1611831
Conc: 1417.61 ng/ml



#41 AR-1268-1

R.T.: 9.383 min
Delta R.T.: 0.000 min
Response: 3035807
Conc: 946.45 ng/ml



#41 AR-1268-1

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 2005487
Conc: 549.21 ng/ml



#42 AR-1268-2

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 2557718
Conc: 830.19 ng/ml

Instrument :
ECD_P
ClientSampleId :



#42 AR-1268-2

R.T.: 8.129 min
Delta R.T.: 0.000 min
Response: 3304939
Conc: 958.02 ng/ml



#43 AR-1268-3

R.T.: 9.700 min
Delta R.T.: 0.000 min
Response: 445521
Conc: 167.58 ng/ml



#43 AR-1268-3

R.T.: 8.335 min
Delta R.T.: 0.001 min
Response: 429938
Conc: 147.68 ng/ml



#44 AR-1268-4

R.T.: 10.135 min
Delta R.T.: 0.000 min
Response: 1780533
Conc: 1461.54 ng/ml

Instrument :
ECD_P
ClientSampleId :



#44 AR-1268-4

R.T.: 8.626 min
Delta R.T.: 0.000 min
Response: 1611831
Conc: 1274.46 ng/ml



#45 AR-1268-5

R.T.: 10.543 min
Delta R.T.: 0.000 min
Response: 2279060
Conc: 239.21 ng/ml



#45 AR-1268-5

R.T.: 8.899 min
Delta R.T.: 0.001 min
Response: 1990105
Conc: 204.77 ng/ml