

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
 Data File : PP038564.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Aug 2021 22:27
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 02:56:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.884	3.871	2089791	1276585	51.782	53.271
2)	SA Decachlor...	10.869	9.130	1342415	1275306	54.562	52.362
Target Compounds							
3)	L1 AR-1016-1	6.202	5.114	660769	461503	509.206	529.455
4)	L1 AR-1016-2	6.226	5.133	945349	656498	511.552	527.880
5)	L1 AR-1016-3	6.291	5.323	594492	352432	520.317	517.564
6)	L1 AR-1016-4	6.399	5.377	485108	274247	520.154	523.689
7)	L1 AR-1016-5	6.713	5.603	433574	338383	511.478	510.184
8)	L2 AR-1221-1	5.130	4.123	75013	45865	160.404	167.640
9)	L2 AR-1221-2	5.233	4.223	107024	66409	302.111	310.425
10)	L2 AR-1221-3	5.320	4.310	410367	259259	385.324	383.963
11)	L3 AR-1232-1	5.320	4.310	410367	259259	414.923	417.400
12)	L3 AR-1232-2	5.907	5.133	512288	656498	1100.145	1150.226
13)	L3 AR-1232-3	6.226	5.323	945349	352432	1107.277	1205.655
14)	L3 AR-1232-4	6.399	5.421	485108	329393	1145.718	1331.321
15)	L3 AR-1232-5	6.496	5.603	368188	338383	1370.190	1251.562
16)	L4 AR-1242-1	6.202	5.114	660769	461503	663.378	668.420
17)	L4 AR-1242-2	6.226	5.133	945349	656498	657.868	672.525
18)	L4 AR-1242-3	6.291	5.323	594492	352432	664.053	670.209
19)	L4 AR-1242-4	6.399	5.421	485108	329393	662.494	689.104
20)	L4 AR-1242-5	7.182	5.981	73342	251736	102.582	406.651 #
21)	L5 AR-1248-1	6.202	5.114	660769	461503	829.268	846.931
22)	L5 AR-1248-2	6.496	5.377	368188	274247	375.531	368.998
23)	L5 AR-1248-3	6.713	5.421	433574	329393	375.609	433.777
24)	L5 AR-1248-4	7.142	5.603	94408	338383	76.452	365.834 #
25)	L5 AR-1248-5	7.182	6.024	73342	42724	59.985	47.319
26)	L6 AR-1254-1	7.111	5.981	306596	251736	268.704	191.684 #
27)	L6 AR-1254-2	7.339	6.142	296521	227109	171.711	197.869
28)	L6 AR-1254-3	7.722	6.578f	229960	446628	128.103	240.025 #
29)	L6 AR-1254-4	8.017	6.798	177460	67663	137.532	57.899 #
30)	L6 AR-1254-5	8.446	7.225	811778	844043	656.347	522.203
31)	L7 AR-1260-1	7.889	6.696	662600	612136	553.644	523.109
32)	L7 AR-1260-2	8.155	6.895	754199	745099	514.044	523.440
33)	L7 AR-1260-3	8.520	7.047	512416	712391	534.046	495.350
34)	L7 AR-1260-4	8.752	7.528	597360	599153	525.318	533.502
35)	L7 AR-1260-5	9.075	7.778	1181235	1436470	526.399	530.321
36)	L8 AR-1262-1	8.520	7.225	512416	844043	344.703	933.802 #
37)	L8 AR-1262-2	9.075	7.778	1181235	1436470	443.671	464.940
38)	L8 AR-1262-3	9.382	8.063	255741	333308	214.641	265.290
39)	L8 AR-1262-4	9.455f	8.125	516396	1094820	640.977	463.243 #
40)	L8 AR-1262-5	10.132	8.625	398162	421643	373.538	370.836
41)	L9 AR-1268-1	9.382	8.063	255741	333308	79.730	91.277

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
 Data File : PP038564.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Aug 2021 22:27
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 02:56:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.455f	8.125	516396	1094820	167.613	317.360 #
44) L9	AR-1268-4	10.132	8.625	398162	421643	326.830	333.389
45) L9	AR-1268-5	10.539	8.897	118081	118680	12.394	12.212



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081821\
Data File : PP038564.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Aug 2021 22:27
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 02:56:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.884 min
Delta R.T.: -0.004 min
Response: 2089791
Conc: 51.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.871 min
Delta R.T.: -0.003 min
Response: 1276585
Conc: 53.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.869 min
Delta R.T.: -0.008 min
Response: 1342415
Conc: 54.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.130 min
Delta R.T.: -0.006 min
Response: 1275306
Conc: 52.36 ng/ml



#3 AR-1016-1


R.T.: 6.202 min
Delta R.T.: -0.004 min
Response: 660769
Conc: 509.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.114 min
Delta R.T.: -0.004 min
Response: 461503
Conc: 529.45 ng/ml




#4 AR-1016-2

R.T.: 6.226 min
Delta R.T.: -0.004 min
Response: 945349
Conc: 511.55 ng/ml



#4 AR-1016-2


R.T.: 5.133 min
Delta R.T.: -0.004 min
Response: 656498
Conc: 527.88 ng/ml



#5 AR-1016-3


R.T.: 6.291 min
Delta R.T.: -0.005 min
Response: 594492
Conc: 520.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.323 min
Delta R.T.: -0.004 min
Response: 352432
Conc: 517.56 ng/ml




#6 AR-1016-4

R.T.: 6.399 min
Delta R.T.: -0.005 min
Response: 485108
Conc: 520.15 ng/ml



#6 AR-1016-4


R.T.: 5.377 min
Delta R.T.: -0.004 min
Response: 274247
Conc: 523.69 ng/ml



#7 AR-1016-5


R.T.: 6.713 min
Delta R.T.: -0.005 min
Response: 433574
Conc: 511.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.603 min
Delta R.T.: -0.004 min
Response: 338383
Conc: 510.18 ng/ml




#8 AR-1221-1

R.T.: 5.130 min
Delta R.T.: -0.010 min
Response: 75013
Conc: 160.40 ng/ml



#8 AR-1221-1


R.T.: 4.123 min
Delta R.T.: -0.005 min
Response: 45865
Conc: 167.64 ng/ml



#9 AR-1221-2


R.T.: 5.233 min
Delta R.T.: -0.005 min
Response: 107024
Conc: 302.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.223 min
Delta R.T.: -0.004 min
Response: 66409
Conc: 310.43 ng/ml




#10 AR-1221-3

R.T.: 5.320 min
Delta R.T.: -0.005 min
Response: 410367
Conc: 385.32 ng/ml



#10 AR-1221-3


R.T.: 4.310 min
Delta R.T.: -0.004 min
Response: 259259
Conc: 383.96 ng/ml



#11 AR-1232-1


R.T.: 5.320 min
Delta R.T.: -0.003 min
Response: 410367
Conc: 414.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.310 min
Delta R.T.: -0.002 min
Response: 259259
Conc: 417.40 ng/ml




#12 AR-1232-2

R.T.: 5.907 min
Delta R.T.: -0.003 min
Response: 512288
Conc: 1100.15 ng/ml



#12 AR-1232-2


R.T.: 5.133 min
Delta R.T.: -0.002 min
Response: 656498
Conc: 1150.23 ng/ml



#13 AR-1232-3


R.T.: 6.226 min
Delta R.T.: -0.003 min
Response: 945349
Conc: 1107.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.323 min
Delta R.T.: -0.002 min
Response: 352432
Conc: 1205.65 ng/ml




#14 AR-1232-4

R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 485108
Conc: 1145.72 ng/ml



#14 AR-1232-4


R.T.: 5.421 min
Delta R.T.: -0.002 min
Response: 329393
Conc: 1331.32 ng/ml



#15 AR-1232-5


R.T.: 6.496 min
Delta R.T.: -0.004 min
Response: 368188
Conc: 1370.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.603 min
Delta R.T.: -0.002 min
Response: 338383
Conc: 1251.56 ng/ml




#16 AR-1242-1

R.T.: 6.202 min
Delta R.T.: -0.003 min
Response: 660769
Conc: 663.38 ng/ml



#16 AR-1242-1


R.T.: 5.114 min
Delta R.T.: -0.002 min
Response: 461503
Conc: 668.42 ng/ml



#17 AR-1242-2


R.T.: 6.226 min
Delta R.T.: -0.003 min
Response: 945349
Conc: 657.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.133 min
Delta R.T.: -0.003 min
Response: 656498
Conc: 672.52 ng/ml




#18 AR-1242-3

R.T.: 6.291 min
Delta R.T.: -0.003 min
Response: 594492
Conc: 664.05 ng/ml



#18 AR-1242-3


R.T.: 5.323 min
Delta R.T.: -0.003 min
Response: 352432
Conc: 670.21 ng/ml



#19 AR-1242-4


R.T.: 6.399 min
Delta R.T.: -0.004 min
Response: 485108
Conc: 662.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.421 min
Delta R.T.: -0.002 min
Response: 329393
Conc: 689.10 ng/ml




#20 AR-1242-5

R.T.: 7.182 min
Delta R.T.: -0.003 min
Response: 73342
Conc: 102.58 ng/ml



#20 AR-1242-5


R.T.: 5.981 min
Delta R.T.: -0.003 min
Response: 251736
Conc: 406.65 ng/ml



#21 AR-1248-1


R.T.: 6.202 min
Delta R.T.: -0.002 min
Response: 660769
Conc: 829.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 461503
Conc: 846.93 ng/ml




#22 AR-1248-2

R.T.: 6.496 min
Delta R.T.: -0.002 min
Response: 368188
Conc: 375.53 ng/ml



#22 AR-1248-2


R.T.: 5.377 min
Delta R.T.: -0.001 min
Response: 274247
Conc: 369.00 ng/ml



#23 AR-1248-3


R.T.: 6.713 min
Delta R.T.: -0.002 min
Response: 433574
Conc: 375.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.421 min
Delta R.T.: 0.000 min
Response: 329393
Conc: 433.78 ng/ml




#24 AR-1248-4

R.T.: 7.142 min
Delta R.T.: -0.002 min
Response: 94408
Conc: 76.45 ng/ml



#24 AR-1248-4


R.T.: 5.603 min
Delta R.T.: -0.002 min
Response: 338383
Conc: 365.83 ng/ml



#25 AR-1248-5


R.T.: 7.182 min
Delta R.T.: -0.002 min
Response: 73342
Conc: 59.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 6.024 min
Delta R.T.: -0.001 min
Response: 42724
Conc: 47.32 ng/ml




#26 AR-1254-1

R.T.: 7.111 min
Delta R.T.: -0.002 min
Response: 306596
Conc: 268.70 ng/ml



#26 AR-1254-1


R.T.: 5.981 min
Delta R.T.: -0.002 min
Response: 251736
Conc: 191.68 ng/ml



#27 AR-1254-2


R.T.: 7.339 min
Delta R.T.: -0.004 min
Response: 296521
Conc: 171.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.142 min
Delta R.T.: -0.002 min
Response: 227109
Conc: 197.87 ng/ml




#28 AR-1254-3

R.T.: 7.722 min
Delta R.T.: -0.003 min
Response: 229960
Conc: 128.10 ng/ml



#28 AR-1254-3


R.T.: 6.578 min
Delta R.T.: 0.018 min
Response: 446628
Conc: 240.03 ng/ml



#29 AR-1254-4


R.T.: 8.017 min
Delta R.T.: -0.005 min
Response: 177460
Conc: 137.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 6.798 min
Delta R.T.: -0.002 min
Response: 67663
Conc: 57.90 ng/ml




#30 AR-1254-5

R.T.: 8.446 min
Delta R.T.: -0.004 min
Response: 811778
Conc: 656.35 ng/ml



#30 AR-1254-5


R.T.: 7.225 min
Delta R.T.: -0.003 min
Response: 844043
Conc: 522.20 ng/ml



#31 AR-1260-1


R.T.: 7.889 min
Delta R.T.: -0.005 min
Response: 662600
Conc: 553.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.696 min
Delta R.T.: -0.004 min
Response: 612136
Conc: 523.11 ng/ml




#32 AR-1260-2

R.T.: 8.155 min
Delta R.T.: -0.005 min
Response: 754199
Conc: 514.04 ng/ml



#32 AR-1260-2


R.T.: 6.895 min
Delta R.T.: -0.005 min
Response: 745099
Conc: 523.44 ng/ml



#33 AR-1260-3


R.T.: 8.520 min
Delta R.T.: -0.006 min
Response: 512416
Conc: 534.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 7.047 min
Delta R.T.: -0.004 min
Response: 712391
Conc: 495.35 ng/ml




#34 AR-1260-4

R.T.: 8.752 min
Delta R.T.: -0.005 min
Response: 597360
Conc: 525.32 ng/ml



#34 AR-1260-4


R.T.: 7.528 min
Delta R.T.: -0.005 min
Response: 599153
Conc: 533.50 ng/ml



#35 AR-1260-5


R.T.: 9.075 min
Delta R.T.: -0.006 min
Response: 1181235
Conc: 526.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.778 min
Delta R.T.: -0.005 min
Response: 1436470
Conc: 530.32 ng/ml




#36 AR-1262-1

R.T.: 8.520 min
Delta R.T.: -0.003 min
Response: 512416
Conc: 344.70 ng/ml



#36 AR-1262-1


R.T.: 7.225 min
Delta R.T.: -0.002 min
Response: 844043
Conc: 933.80 ng/ml



#37 AR-1262-2


R.T.: 9.075 min
Delta R.T.: -0.002 min
Response: 1181235
Conc: 443.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.778 min
Delta R.T.: -0.001 min
Response: 1436470
Conc: 464.94 ng/ml




#38 AR-1262-3

R.T.: 9.382 min
Delta R.T.: 0.000 min
Response: 255741
Conc: 214.64 ng/ml



#38 AR-1262-3


R.T.: 8.063 min
Delta R.T.: -0.002 min
Response: 333308
Conc: 265.29 ng/ml



#39 AR-1262-4


R.T.: 9.455 min
Delta R.T.: -0.020 min
Response: 516396
Conc: 640.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.125 min
Delta R.T.: -0.003 min
Response: 1094820
Conc: 463.24 ng/ml




#40 AR-1262-5

R.T.: 10.132 min
Delta R.T.: -0.003 min
Response: 398162
Conc: 373.54 ng/ml



#40 AR-1262-5


R.T.: 8.625 min
Delta R.T.: -0.001 min
Response: 421643
Conc: 370.84 ng/ml



#41 AR-1268-1


R.T.: 9.382 min
Delta R.T.: 0.000 min
Response: 255741
Conc: 79.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 333308
Conc: 91.28 ng/ml




#42 AR-1268-2

R.T.: 9.455 min
Delta R.T.: -0.022 min
Response: 516396
Conc: 167.61 ng/ml



#42 AR-1268-2


R.T.: 8.125 min
Delta R.T.: -0.004 min
Response: 1094820
Conc: 317.36 ng/ml



#44 AR-1268-4


R.T.: 10.132 min
Delta R.T.: -0.002 min
Response: 398162
Conc: 326.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#44 AR-1268-4

R.T.: 8.625 min
Delta R.T.: 0.000 min
Response: 421643
Conc: 333.39 ng/ml



#45 AR-1268-5

R.T.: 10.539 min
Delta R.T.: -0.004 min
Response: 118081
Conc: 12.39 ng/ml



#45 AR-1268-5

R.T.: 8.897 min
Delta R.T.: 0.000 min
Response: 118680
Conc: 12.21 ng/ml