

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
 Data File : PP038611.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Aug 2021 21:20
 Operator : AJ\MA
 Sample : M3465-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 01:40:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.884	3.870	568606	333292	14.089	13.908
2)	SA Decachlor...	10.871	9.131	412792	412571	16.778	16.939
Target Compounds							
8)	L2 AR-1221-1	5.139	0.000	16235	0	34.717	N.D. #
28)	L6 AR-1254-3	7.745f	0.000	21116	0	11.763	N.D. #
32)	L7 AR-1260-2	8.134f	0.000	31033	0	21.151	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
Data File : PP038611.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Aug 2021 21:20
Operator : AJ\MA
Sample : M3465-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 01:40:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.884 min
Delta R.T.: -0.004 min
Response: 568606
Conc: 14.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.870 min
Delta R.T.: -0.004 min
Response: 333292
Conc: 13.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.871 min
Delta R.T.: -0.006 min
Response: 412792
Conc: 16.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.131 min
Delta R.T.: -0.005 min
Response: 412571
Conc: 16.94 ng/ml



#8 AR-1221-1


R.T.: 5.139 min
Delta R.T.: -0.001 min
Response: 16235
Conc: 34.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.129 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.745 min
Delta R.T.: 0.020 min
Response: 21116
Conc: 11.76 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.560 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 8.134 min
Delta R.T.: -0.026 min
Response: 31033
Conc: 21.15 ng/ml

Instrument :
ECD_P
ClientSampleId :



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.899 min
Response: 0
Conc: N.D.