

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
 Data File : PP038586.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Aug 2021 12:20
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 01:35:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.885	3.870	2084128	1313037	51.641	54.792
2) SA Decachlor...	10.872	9.131	1355085	1255433	55.076	51.546
Target Compounds						
26) L6 AR-1254-1	7.113	5.982	572620	659463	501.850	502.147
27) L6 AR-1254-2	7.343	6.142	805164	571174	466.260	497.636
28) L6 AR-1254-3	7.724	6.559	864263	951916	481.451	511.576
29) L6 AR-1254-4	8.021	6.799	602872	582888	467.228	498.780
30) L6 AR-1254-5	8.449	7.227	606659	798627	490.502	494.105



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
Data File : PP038586.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Aug 2021 12:20
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 01:35:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.885 min
Delta R.T.: -0.003 min
Response: 2084128
Conc: 51.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.870 min
Delta R.T.: -0.004 min
Response: 1313037
Conc: 54.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.872 min
Delta R.T.: -0.005 min
Response: 1355085
Conc: 55.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.131 min
Delta R.T.: -0.004 min
Response: 1255433
Conc: 51.55 ng/ml



#26 AR-1254-1


R.T.: 7.113 min
Delta R.T.: 0.000 min
Response: 572620
Conc: 501.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.982 min
Delta R.T.: 0.000 min
Response: 659463
Conc: 502.15 ng/ml




#27 AR-1254-2

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 805164
Conc: 466.26 ng/ml



#27 AR-1254-2


R.T.: 6.142 min
Delta R.T.: -0.002 min
Response: 571174
Conc: 497.64 ng/ml



#28 AR-1254-3


R.T.: 7.724 min
Delta R.T.: 0.000 min
Response: 864263
Conc: 481.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.559 min
Delta R.T.: -0.001 min
Response: 951916
Conc: 511.58 ng/ml




#29 AR-1254-4

R.T.: 8.021 min
Delta R.T.: 0.000 min
Response: 602872
Conc: 467.23 ng/ml



#29 AR-1254-4


R.T.: 6.799 min
Delta R.T.: -0.001 min
Response: 582888
Conc: 498.78 ng/ml



#30 AR-1254-5

R.T.: 8.449 min
Delta R.T.: -0.001 min
Response: 606659
Conc: 490.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.227 min
Delta R.T.: 0.000 min
Response: 798627
Conc: 494.10 ng/ml