

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
 Data File : PP038595.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Aug 2021 16:47
 Operator : AJ\MA
 Sample : M3449-13
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 01:37:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.886	3.871	799995	494866	19.823	20.650
2)	SA Decachlor...	10.873	9.133	467269	461227	18.992	18.937

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081921\
Data File : PP038595.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Aug 2021 16:47
Operator : AJ\MA
Sample : M3449-13
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 01:37:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.886 min
Delta R.T.: -0.002 min
Response: 799995
Conc: 19.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP5




#1 Tetrachloro-m-xylene

R.T.: 3.871 min
Delta R.T.: -0.003 min
Response: 494866
Conc: 20.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.873 min
Delta R.T.: -0.004 min
Response: 467269
Conc: 18.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.133 min
Delta R.T.: -0.003 min
Response: 461227
Conc: 18.94 ng/ml