

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
 Data File : PP050731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2022 15:28
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 04:02:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.592	75548951	150.4E6	46.935	48.005
2) SA Decachlor...	10.099	8.555	75179349	69437301	42.694	45.116
Target Compounds						
16) L4 AR-1242-1	5.526	4.663	23819654	38931681	469.983	469.623
17) L4 AR-1242-2	5.548	4.681	34010458	57165253	470.801	489.513
18) L4 AR-1242-3	5.609	4.856	20240032	29376609	457.677	479.066
19) L4 AR-1242-4	5.711	4.938	17731986	27766705	475.845	474.324
20) L4 AR-1242-5	6.450	5.458	19858917	34344455	470.575	494.095



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
Data File : PP050731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2022 15:28
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 04:02:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.349 min
Delta R.T.: 0.003 min
Response: 75548951
Conc: 46.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: 0.000 min
Response: 150363265
Conc: 48.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.099 min
Delta R.T.: 0.009 min
Response: 75179349
Conc: 42.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 69437301
Conc: 45.12 ng/ml



#16 AR-1242-1


R.T.: 5.526 min
Delta R.T.: 0.004 min
Response: 23819654
Conc: 469.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.663 min
Delta R.T.: -0.002 min
Response: 38931681
Conc: 469.62 ng/ml




#17 AR-1242-2

R.T.: 5.548 min
Delta R.T.: 0.003 min
Response: 34010458
Conc: 470.80 ng/ml



#17 AR-1242-2


R.T.: 4.681 min
Delta R.T.: -0.002 min
Response: 57165253
Conc: 489.51 ng/ml



#18 AR-1242-3


R.T.: 5.609 min
Delta R.T.: 0.003 min
Response: 20240032
Conc: 457.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.856 min
Delta R.T.: -0.002 min
Response: 29376609
Conc: 479.07 ng/ml




#19 AR-1242-4

R.T.: 5.711 min
Delta R.T.: 0.004 min
Response: 17731986
Conc: 475.84 ng/ml



#19 AR-1242-4


R.T.: 4.938 min
Delta R.T.: -0.002 min
Response: 27766705
Conc: 474.32 ng/ml



#20 AR-1242-5

R.T.: 6.450 min
Delta R.T.: 0.004 min
Response: 19858917
Conc: 470.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.458 min
Delta R.T.: -0.002 min
Response: 34344455
Conc: 494.10 ng/ml