

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
 Data File : PP050733.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2022 16:01
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/22/2022
 Supervised By :Ankita Jodhani 08/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 04:03:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.592	76561063	154.5E6	47.563	49.335
2) SA Decachlor...	10.096	8.555	77998075	71145517	44.294	46.226
Target Compounds						
26) L6 AR-1254-1	6.379	5.458	33975488	69758760	459.220	490.840
27) L6 AR-1254-2	6.598	5.605	52217521	59634188	449.485	503.899
28) L6 AR-1254-3	6.969	6.006	53748358	86056358	433.754	500.120
29) L6 AR-1254-4	7.256	6.234	42638883	58046225	435.780	550.161 #
30) L6 AR-1254-5	7.676	6.649	42478090	70826420	423.406m	531.914 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
Data File : PP050733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2022 16:01
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 04:03:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 76561063
Conc: 47.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: 0.000 min
Response: 154528486
Conc: 49.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.096 min
Delta R.T.: 0.006 min
Response: 77998075
Conc: 44.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 71145517
Conc: 46.23 ng/ml




#26 AR-1254-1

R.T.: 6.379 min
Delta R.T.: 0.000 min
Response: 33975488
Conc: 459.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022




#26 AR-1254-1

R.T.: 5.458 min
Delta R.T.: -0.002 min
Response: 69758760
Conc: 490.84 ng/ml




#27 AR-1254-2

R.T.: 6.598 min
Delta R.T.: 0.000 min
Response: 52217521
Conc: 449.49 ng/ml



#27 AR-1254-2

R.T.: 5.605 min
Delta R.T.: -0.002 min
Response: 59634188
Conc: 503.90 ng/ml




#28 AR-1254-3

R.T.: 6.969 min
Delta R.T.: 0.001 min
Response: 53748358
Conc: 433.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022




#28 AR-1254-3

R.T.: 6.006 min
Delta R.T.: -0.002 min
Response: 86056358
Conc: 500.12 ng/ml




#29 AR-1254-4

R.T.: 7.256 min
Delta R.T.: 0.000 min
Response: 42638883
Conc: 435.78 ng/ml



#29 AR-1254-4

R.T.: 6.234 min
Delta R.T.: -0.002 min
Response: 58046225
Conc: 550.16 ng/ml




#30 AR-1254-5

R.T.: 7.676 min
Delta R.T.: 0.002 min
Response: 42478090
Conc: 423.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022



#30 AR-1254-5

R.T.: 6.649 min
Delta R.T.: -0.002 min
Response: 70826420
Conc: 531.91 ng/ml