

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
 Data File : PP050737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2022 17:07
 Operator : YP\AJ
 Sample : N4256-02DL 10X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 34090DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 04:07:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.592	5380196	9472100	3.342	3.024
2) SA Decachlor...	10.097	8.553	21261422	19556340	12.074	12.706
Target Compounds						
16) L4 AR-1242-1	5.523	4.663	35079949	52034888	692.159	627.684
17) L4 AR-1242-2	5.545	4.681	49134131	75050826	680.155	642.670
18) L4 AR-1242-3	5.607	4.856	30634123	43601209	692.713	711.037
19) L4 AR-1242-4	5.708	4.938	29177128	66483054	782.980	1135.694 #
20) L4 AR-1242-5	6.447	5.458	21576594	46722359	511.277	672.169 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
Data File : PP050737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2022 17:07
Operator : YP\AJ
Sample : N4256-02DL 10X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
34090DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 04:07:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 5380196
Conc: 3.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
34090DL




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: 0.000 min
Response: 9472100
Conc: 3.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: 0.007 min
Response: 21261422
Conc: 12.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.553 min
Delta R.T.: -0.002 min
Response: 19556340
Conc: 12.71 ng/ml



#16 AR-1242-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 35079949
Conc: 692.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
34090DL




#16 AR-1242-1

R.T.: 4.663 min
Delta R.T.: -0.002 min
Response: 52034888
Conc: 627.68 ng/ml




#17 AR-1242-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 49134131
Conc: 680.16 ng/ml



#17 AR-1242-2


R.T.: 4.681 min
Delta R.T.: -0.002 min
Response: 75050826
Conc: 642.67 ng/ml



#18 AR-1242-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 30634123
Conc: 692.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
34090DL




#18 AR-1242-3

R.T.: 4.856 min
Delta R.T.: -0.002 min
Response: 43601209
Conc: 711.04 ng/ml




#19 AR-1242-4

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 29177128
Conc: 782.98 ng/ml



#19 AR-1242-4


R.T.: 4.938 min
Delta R.T.: -0.002 min
Response: 66483054
Conc: 1135.69 ng/ml



#20 AR-1242-5

R.T.: 6.447 min
Delta R.T.: 0.002 min
Response: 21576594
Conc: 511.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
34090DL



#20 AR-1242-5

R.T.: 5.458 min
Delta R.T.: -0.002 min
Response: 46722359
Conc: 672.17 ng/ml