

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
 Data File : PP050740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2022 17:57
 Operator : YP\AJ
 Sample : N4259-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 8218

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 04:09:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.597	24312187	40871872	15.104	13.049
2) SA Decachlor...	10.094	8.553	35699432	31261284	20.273	20.312

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081922\
Data File : PP050740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2022 17:57
Operator : YP\AJ
Sample : N4259-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
8218

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 04:09:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: 0.004 min
Response: 24312187
Conc: 15.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
8218




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.004 min
Response: 40871872
Conc: 13.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.004 min
Response: 35699432
Conc: 20.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.553 min
Delta R.T.: -0.002 min
Response: 31261284
Conc: 20.31 ng/ml