

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082021\
 Data File : PP038688.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Aug 2021 2:51
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 02:05:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.884	3.870	2010620	1269729	49.820	52.985
2) SA Decachlor...	10.871	9.129	1304081	1248253	53.003	51.251
Target Compounds						
16) L4 AR-1242-1	6.202	5.113	469960	335909	471.815	486.516
17) L4 AR-1242-2	6.226	5.132	674687	476849	469.514	488.490
18) L4 AR-1242-3	6.292	5.321	423303	254970	472.833	484.868
19) L4 AR-1242-4	6.400	5.420	347054	238593	473.959	499.146
20) L4 AR-1242-5	7.182	5.980	327993	309528	458.758	500.006



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082021\
Data File : PP038688.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Aug 2021 2:51
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 02:05:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.884 min
Delta R.T.: -0.004 min
Response: 2010620
Conc: 49.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.870 min
Delta R.T.: -0.004 min
Response: 1269729
Conc: 52.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.871 min
Delta R.T.: -0.006 min
Response: 1304081
Conc: 53.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.129 min
Delta R.T.: -0.007 min
Response: 1248253
Conc: 51.25 ng/ml



#16 AR-1242-1


R.T.: 6.202 min
Delta R.T.: -0.003 min
Response: 469960
Conc: 471.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.113 min
Delta R.T.: -0.003 min
Response: 335909
Conc: 486.52 ng/ml




#17 AR-1242-2

R.T.: 6.226 min
Delta R.T.: -0.003 min
Response: 674687
Conc: 469.51 ng/ml



#17 AR-1242-2


R.T.: 5.132 min
Delta R.T.: -0.004 min
Response: 476849
Conc: 488.49 ng/ml



#18 AR-1242-3


R.T.: 6.292 min
Delta R.T.: -0.003 min
Response: 423303
Conc: 472.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.321 min
Delta R.T.: -0.004 min
Response: 254970
Conc: 484.87 ng/ml




#19 AR-1242-4

R.T.: 6.400 min
Delta R.T.: -0.003 min
Response: 347054
Conc: 473.96 ng/ml



#19 AR-1242-4


R.T.: 5.420 min
Delta R.T.: -0.004 min
Response: 238593
Conc: 499.15 ng/ml



#20 AR-1242-5

R.T.: 7.182 min
Delta R.T.: -0.003 min
Response: 327993
Conc: 458.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.980 min
Delta R.T.: -0.005 min
Response: 309528
Conc: 500.01 ng/ml