

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082024\
 Data File : PP066710.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 00:50
 Operator : YP\AJ
 Sample : P3390-07
 Misc : AR1660 LOD 25 PPB
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2024

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :Ankita Jodhani 08/21/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 05:53:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 21 05:19:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.703	3.978	24068663	20425402	16.254	15.622
2) SA Decachlor...	10.622	9.142	14121234	12422484	16.092	14.912
Target Compounds						
3) L1 AR-1016-1	5.876	5.085	929941	824799	21.819m	20.575
4) L1 AR-1016-2	5.900	5.105	1403270	1200525	20.976	20.146
5) L1 AR-1016-3	5.962	5.287	941073	627438	22.234	20.990
6) L1 AR-1016-4	6.061	5.325	610342	666568	18.706	25.436 #
7) L1 AR-1016-5	6.356	5.546	672476	699427	21.204m	21.539
31) L7 AR-1260-1	7.485	6.597	1233039	1391609	22.003	23.544
32) L7 AR-1260-2	7.740	6.785	1371684	1556959	21.546	22.850
33) L7 AR-1260-3	8.102	6.944	1352736	1507892	30.418	23.208
34) L7 AR-1260-4	8.338	7.421	946977	1168589	17.942m	25.948 #
35) L7 AR-1260-5	8.675	7.661	1806619	1890450	20.310	19.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082024\
Data File : PP066710.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 00:50
Operator : YP\AJ
Sample : P3390-07
Misc : AR1660 LOD 25 PPB
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 05:53:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 21 05:19:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

