

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082024\
 Data File : PP066695.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2024 20:39
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP082024

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :Ankita Jodhani 08/21/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 01:42:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 21 01:33:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.702	3.979	75429903	66371601	50.196	49.877
2) SA Decachlor...	10.623	9.144	77432504	70850854	50.367	49.501
Target Compounds						
41) L9 AR-1268-1	9.006	7.949	65666527	65051439	499.249m	492.094m
42) L9 AR-1268-2	9.108	8.015	58158890	59661647	499.772	491.695
43) L9 AR-1268-3	9.354	8.230	53176732	51849774	507.722	489.050
44) L9 AR-1268-4	9.807	8.539	15745977	18016545	484.956	517.274
45) L9 AR-1268-5	10.255	8.859	149.8E6	153.5E6	503.641	498.209

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082024\
Data File : PP066695.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2024 20:39
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP082024


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 01:42:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 21 01:33:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.702 min
Delta R.T.: -0.002 min
Response: 75429903
Conc: 50.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP082024


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024




#1 Tetrachloro-m-xylene

R.T.: 3.979 min
Delta R.T.: 0.000 min
Response: 66371601
Conc: 49.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.623 min
Delta R.T.: -0.003 min
Response: 77432504
Conc: 50.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.144 min
Delta R.T.: 0.000 min
Response: 70850854
Conc: 49.50 ng/ml




#41 AR-1268-1

R.T.: 9.006 min
Delta R.T.: -0.003 min
Response: 65666527
Conc: 499.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP082024


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024




#41 AR-1268-1

R.T.: 7.949 min
Delta R.T.: 0.000 min
Response: 65051439
Conc: 492.09 ng/ml




#42 AR-1268-2

R.T.: 9.108 min
Delta R.T.: -0.002 min
Response: 58158890
Conc: 499.77 ng/ml



#42 AR-1268-2

R.T.: 8.015 min
Delta R.T.: 0.000 min
Response: 59661647
Conc: 491.69 ng/ml




#43 AR-1268-3

R.T.: 9.354 min
Delta R.T.: -0.001 min
Response: 53176732
Conc: 507.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP082024


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024




#43 AR-1268-3

R.T.: 8.230 min
Delta R.T.: 0.000 min
Response: 51849774
Conc: 489.05 ng/ml




#44 AR-1268-4

R.T.: 9.807 min
Delta R.T.: -0.003 min
Response: 15745977
Conc: 484.96 ng/ml



#44 AR-1268-4

R.T.: 8.539 min
Delta R.T.: 0.001 min
Response: 18016545
Conc: 517.27 ng/ml




#45 AR-1268-5

R.T.: 10.255 min
Delta R.T.: -0.003 min
Response: 149836232
Conc: 503.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP082024

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :Ankita Jodhani 08/21/2024



#45 AR-1268-5

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 153495815
Conc: 498.21 ng/ml