

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD P\Data\PP082118\
 Data File : PP027137.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Aug 2018 22:30
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 8/22/2018 1:07:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 03:27:20 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073018.M
 Quant Title : 8080.M
 QLast Update : Tue Jul 31 01:31:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.044	12.156	181.0E6	317.0E6	757.861	782.736
Target Compounds							
1) T	Dalapon	3.311	2.983	374.4E6	377.9E6	637.279	672.561
2) T	3,5-DICHL...	10.458	10.248	307.2E6	390.1E6	667.462	729.194
3) T	4-Nitroph...	11.617	11.297	293.8E6	180.6E6	640.570	745.620
5) T	DICAMBA	12.366	12.468	980.0E6	1190.3E6	693.740	731.456
6) T	MCPD	12.711	12.662	30454875	68290146	78.498	79.242
7) T	MCPA	12.954	13.049	49496019	101.7E6	73.760	80.864m
8) T	DICHLORPROP	13.561	13.582	209.9E6	290.4E6	664.108m	748.252
9) T	2,4-D	13.911	14.065	276.1E6	398.0E6	706.771	754.655
10) T	Pentachlo...	14.330	14.679	5290.3E6	4832.9E6	695.092	724.509
11) T	2,4,5-TP ...	15.192	15.249	2003.6E6	1942.7E6	698.332	748.605
12) T	2,4,5-T	15.579	15.886	1798.7E6	1669.6E6	704.880	766.349
13) T	2,4-DB	16.341	16.489	160.5E6	218.0E6	733.932	765.159
14) T	DINOSEB	17.833	16.885	1366.5E6	1328.4E6	675.674	736.014
15) T	Picloram	17.588	18.257	3073.8E6	2801.0E6	744.264	812.134
16) T	DCPA	18.197	18.144	2783.3E6	2384.3E6	693.852	740.808

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD P\Data\PP082118\
Data File : PP027137.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Aug 2018 22:30
Operator : EI\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
8/22/2018 1:07:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 03:27:20 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073018.M
Quant Title : 8080.M
QLast Update : Tue Jul 31 01:31:14 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

