

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082124\
 Data File : PP066743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 10:18
 Operator : YP\AJ
 Sample : P3390-08
 Misc : AR1660 LOQ 50 PPB
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOQ-WATER-02-QT3-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 04:07:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 21 05:19:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.704	3.981	25586917	20064631	17.279	15.346
2)	SA Decachlor...	10.626	9.146	16565429	13123113	18.877	15.753
Target Compounds							
3)	L1 AR-1016-1	5.878	5.088	1978822	1610454	46.430	40.173
4)	L1 AR-1016-2	5.900	5.108	3132420	2649300	46.823	44.458
5)	L1 AR-1016-3	5.964	5.290	1970700	1108964	46.559	37.099
6)	L1 AR-1016-4	6.063	5.330	1413448	1034055	43.321	39.459
7)	L1 AR-1016-5	6.358	5.549	1516926	1301273	47.830	40.073
8)	L2 AR-1221-1	4.906	0.000	251665	0	14.871	N.D. #
9)	L2 AR-1221-2	5.013	4.290	513958	491642	41.448	41.889
10)	L2 AR-1221-3	5.076	4.363	1188455	1057452	31.317	29.016
11)	L3 AR-1232-1	5.076	4.363	1188455	1057452	37.370	34.507
12)	L3 AR-1232-2	5.609	5.108	1610622	2649300	92.987	98.188
13)	L3 AR-1232-3	5.900	5.290	3132420	1108964	103.764	82.690
14)	L3 AR-1232-4	6.063	5.373	1413448	1164946	95.970	90.627
15)	L3 AR-1232-5	6.151	5.549	1272314	1301273	107.184	93.962
16)	L4 AR-1242-1	5.878	5.088	1978822	1610454	61.364	52.490
17)	L4 AR-1242-2	5.900	5.108	3132420	2649300	61.896	58.468
18)	L4 AR-1242-3	5.964	5.290	1970700	1108964	62.367	47.475
19)	L4 AR-1242-4	6.063	5.373	1413448	1164946	57.039	46.296
20)	L4 AR-1242-5	6.801	5.908	353224	1250872	14.057	44.222 #
21)	L5 AR-1248-1	5.878	5.088	1978822	1610454	77.423	66.516
22)	L5 AR-1248-2	6.151	5.330	1272314	1034055	31.665	28.631
23)	L5 AR-1248-3	6.358	5.373	1516926	1164946	34.790	31.203
24)	L5 AR-1248-4	6.760	5.549	337018	1301273	7.629	29.492 #
25)	L5 AR-1248-5	6.801	5.952	353224	261780	7.705	6.930
26)	L6 AR-1254-1	6.736	5.908	1349384	1250872	27.334	20.880
27)	L6 AR-1254-2	6.953	6.056	1352657	870960	18.540	16.298
28)	L6 AR-1254-3	7.324	6.482	541126	1836336	7.745	23.223 #
29)	L6 AR-1254-4	7.616	0.000	425395	0	9.555	N.D. #
30)	L6 AR-1254-5	8.026	7.121	4666712	2407238	70.392	37.452 #
31)	L7 AR-1260-1	7.486	6.600	2576390	2479785	45.975	41.955
32)	L7 AR-1260-2	7.741	6.788	3120493	2742962	49.015	40.256
33)	L7 AR-1260-3	8.102	6.947	3111616	2744074	69.969	42.233 #
34)	L7 AR-1260-4	8.340	7.423	3947480	2077172	74.792	46.123 #
35)	L7 AR-1260-5	8.677	7.664	5236244	3830965	58.866	39.698 #
36)	L8 AR-1262-1	8.026	7.121	4666712	2407238	103.358	75.310 #
37)	L8 AR-1262-2	8.677	7.423	5236244	2077172	47.952	33.313 #
38)	L8 AR-1262-3	9.026	7.950	2340550	959288	32.969	21.438 #
39)	L8 AR-1262-4	9.100	8.016	1468864	2752843	25.169	33.204 #
40)	L8 AR-1262-5	9.812	8.542	774119	713876	23.186	22.544
41)	L9 AR-1268-1	9.026f	7.950	2340550	959288	18.675	7.662 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082124\
Data File : PP066743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 10:18
Operator : YP\AJ
Sample : P3390-08
Misc : AR1660 LOQ 50 PPB
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 04:07:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 21 05:19:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.100	8.016	1468864	2752843	13.317	23.769 #
43)	L9 AR-1268-3	9.357	0.000	69169	0	0.697	N.D. #
44)	L9 AR-1268-4	9.812	8.542	774119	713876	25.406	21.096
45)	L9 AR-1268-5	10.255	8.862	270114	211739	0.947	0.729



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082124\
Data File : PP066743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 10:18
Operator : YP\AJ
Sample : P3390-08
Misc : AR1660 LOQ 50 PPB
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 04:07:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 21 05:19:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 25586917
Conc: 17.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: 0.001 min
Response: 20064631
Conc: 15.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.626 min
Delta R.T.: 0.000 min
Response: 16565429
Conc: 18.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.146 min
Delta R.T.: 0.000 min
Response: 13123113
Conc: 15.75 ng/ml



#3 AR-1016-1


R.T.: 5.878 min
Delta R.T.: -0.001 min
Response: 1978822
Conc: 46.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#3 AR-1016-1

R.T.: 5.088 min
Delta R.T.: 0.002 min
Response: 1610454
Conc: 40.17 ng/ml




#4 AR-1016-2

R.T.: 5.900 min
Delta R.T.: -0.001 min
Response: 3132420
Conc: 46.82 ng/ml



#4 AR-1016-2


R.T.: 5.108 min
Delta R.T.: 0.002 min
Response: 2649300
Conc: 44.46 ng/ml



#5 AR-1016-3


R.T.: 5.964 min
Delta R.T.: 0.000 min
Response: 1970700
Conc: 46.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#5 AR-1016-3

R.T.: 5.290 min
Delta R.T.: 0.001 min
Response: 1108964
Conc: 37.10 ng/ml




#6 AR-1016-4

R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 1413448
Conc: 43.32 ng/ml



#6 AR-1016-4


R.T.: 5.330 min
Delta R.T.: 0.003 min
Response: 1034055
Conc: 39.46 ng/ml



#7 AR-1016-5


R.T.: 6.358 min
Delta R.T.: -0.001 min
Response: 1516926
Conc: 47.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#7 AR-1016-5

R.T.: 5.549 min
Delta R.T.: 0.002 min
Response: 1301273
Conc: 40.07 ng/ml




#8 AR-1221-1

R.T.: 4.906 min
Delta R.T.: -0.005 min
Response: 251665
Conc: 14.87 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.195 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.013 min
Delta R.T.: 0.015 min
Response: 513958
Conc: 41.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#9 AR-1221-2

R.T.: 4.290 min
Delta R.T.: 0.006 min
Response: 491642
Conc: 41.89 ng/ml




#10 AR-1221-3

R.T.: 5.076 min
Delta R.T.: 0.000 min
Response: 1188455
Conc: 31.32 ng/ml



#10 AR-1221-3


R.T.: 4.363 min
Delta R.T.: 0.002 min
Response: 1057452
Conc: 29.02 ng/ml



#11 AR-1232-1


R.T.: 5.076 min
Delta R.T.: 0.001 min
Response: 1188455
Conc: 37.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#11 AR-1232-1

R.T.: 4.363 min
Delta R.T.: 0.003 min
Response: 1057452
Conc: 34.51 ng/ml




#12 AR-1232-2

R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 1610622
Conc: 92.99 ng/ml



#12 AR-1232-2


R.T.: 5.108 min
Delta R.T.: 0.003 min
Response: 2649300
Conc: 98.19 ng/ml



#13 AR-1232-3


R.T.: 5.900 min
Delta R.T.: 0.000 min
Response: 3132420
Conc: 103.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#13 AR-1232-3

R.T.: 5.290 min
Delta R.T.: 0.004 min
Response: 1108964
Conc: 82.69 ng/ml




#14 AR-1232-4

R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 1413448
Conc: 95.97 ng/ml



#14 AR-1232-4


R.T.: 5.373 min
Delta R.T.: 0.004 min
Response: 1164946
Conc: 90.63 ng/ml



#15 AR-1232-5


R.T.: 6.151 min
Delta R.T.: 0.000 min
Response: 1272314
Conc: 107.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#15 AR-1232-5

R.T.: 5.549 min
Delta R.T.: 0.003 min
Response: 1301273
Conc: 93.96 ng/ml




#16 AR-1242-1

R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 1978822
Conc: 61.36 ng/ml



#16 AR-1242-1


R.T.: 5.088 min
Delta R.T.: 0.002 min
Response: 1610454
Conc: 52.49 ng/ml



#17 AR-1242-2


R.T.: 5.900 min
Delta R.T.: 0.000 min
Response: 3132420
Conc: 61.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#17 AR-1242-2

R.T.: 5.108 min
Delta R.T.: 0.002 min
Response: 2649300
Conc: 58.47 ng/ml




#18 AR-1242-3

R.T.: 5.964 min
Delta R.T.: 0.000 min
Response: 1970700
Conc: 62.37 ng/ml



#18 AR-1242-3


R.T.: 5.290 min
Delta R.T.: 0.002 min
Response: 1108964
Conc: 47.48 ng/ml



#19 AR-1242-4


R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 1413448
Conc: 57.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#19 AR-1242-4

R.T.: 5.373 min
Delta R.T.: 0.002 min
Response: 1164946
Conc: 46.30 ng/ml




#20 AR-1242-5

R.T.: 6.801 min
Delta R.T.: -0.001 min
Response: 353224
Conc: 14.06 ng/ml



#20 AR-1242-5


R.T.: 5.908 min
Delta R.T.: 0.003 min
Response: 1250872
Conc: 44.22 ng/ml



#21 AR-1248-1


R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 1978822
Conc: 77.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#21 AR-1248-1

R.T.: 5.088 min
Delta R.T.: 0.002 min
Response: 1610454
Conc: 66.52 ng/ml




#22 AR-1248-2

R.T.: 6.151 min
Delta R.T.: 0.000 min
Response: 1272314
Conc: 31.66 ng/ml



#22 AR-1248-2


R.T.: 5.330 min
Delta R.T.: 0.003 min
Response: 1034055
Conc: 28.63 ng/ml



#23 AR-1248-3


R.T.: 6.358 min
Delta R.T.: 0.000 min
Response: 1516926
Conc: 34.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#23 AR-1248-3

R.T.: 5.373 min
Delta R.T.: 0.002 min
Response: 1164946
Conc: 31.20 ng/ml




#24 AR-1248-4

R.T.: 6.760 min
Delta R.T.: -0.001 min
Response: 337018
Conc: 7.63 ng/ml



#24 AR-1248-4


R.T.: 5.549 min
Delta R.T.: 0.002 min
Response: 1301273
Conc: 29.49 ng/ml



#25 AR-1248-5


R.T.: 6.801 min
Delta R.T.: 0.000 min
Response: 353224
Conc: 7.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#25 AR-1248-5

R.T.: 5.952 min
Delta R.T.: 0.005 min
Response: 261780
Conc: 6.93 ng/ml




#26 AR-1254-1

R.T.: 6.736 min
Delta R.T.: -0.002 min
Response: 1349384
Conc: 27.33 ng/ml



#26 AR-1254-1


R.T.: 5.908 min
Delta R.T.: 0.000 min
Response: 1250872
Conc: 20.88 ng/ml



#27 AR-1254-2


R.T.: 6.953 min
Delta R.T.: -0.002 min
Response: 1352657
Conc: 18.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#27 AR-1254-2

R.T.: 6.056 min
Delta R.T.: 0.000 min
Response: 870960
Conc: 16.30 ng/ml




#28 AR-1254-3

R.T.: 7.324 min
Delta R.T.: 0.000 min
Response: 541126
Conc: 7.74 ng/ml



#28 AR-1254-3


R.T.: 6.482 min
Delta R.T.: 0.014 min
Response: 1836336
Conc: 23.22 ng/ml



#29 AR-1254-4


R.T.: 7.616 min
Delta R.T.: 0.007 min
Response: 425395
Conc: 9.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.697 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.026 min
Delta R.T.: -0.003 min
Response: 4666712
Conc: 70.39 ng/ml



#30 AR-1254-5


R.T.: 7.121 min
Delta R.T.: 0.000 min
Response: 2407238
Conc: 37.45 ng/ml



#31 AR-1260-1


R.T.: 7.486 min
Delta R.T.: -0.002 min
Response: 2576390
Conc: 45.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#31 AR-1260-1

R.T.: 6.600 min
Delta R.T.: 0.001 min
Response: 2479785
Conc: 41.96 ng/ml




#32 AR-1260-2

R.T.: 7.741 min
Delta R.T.: -0.001 min
Response: 3120493
Conc: 49.02 ng/ml



#32 AR-1260-2


R.T.: 6.788 min
Delta R.T.: 0.000 min
Response: 2742962
Conc: 40.26 ng/ml



#33 AR-1260-3


R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 3111616
Conc: 69.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#33 AR-1260-3

R.T.: 6.947 min
Delta R.T.: 0.001 min
Response: 2744074
Conc: 42.23 ng/ml




#34 AR-1260-4

R.T.: 8.340 min
Delta R.T.: 0.000 min
Response: 3947480
Conc: 74.79 ng/ml



#34 AR-1260-4


R.T.: 7.423 min
Delta R.T.: 0.000 min
Response: 2077172
Conc: 46.12 ng/ml



#35 AR-1260-5


R.T.: 8.677 min
Delta R.T.: -0.003 min
Response: 5236244
Conc: 58.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#35 AR-1260-5

R.T.: 7.664 min
Delta R.T.: 0.000 min
Response: 3830965
Conc: 39.70 ng/ml




#36 AR-1262-1

R.T.: 8.026 min
Delta R.T.: 0.002 min
Response: 4666712
Conc: 103.36 ng/ml



#36 AR-1262-1


R.T.: 7.121 min
Delta R.T.: 0.002 min
Response: 2407238
Conc: 75.31 ng/ml



#37 AR-1262-2


R.T.: 8.677 min
Delta R.T.: -0.001 min
Response: 5236244
Conc: 47.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#37 AR-1262-2

R.T.: 7.423 min
Delta R.T.: 0.002 min
Response: 2077172
Conc: 33.31 ng/ml




#38 AR-1262-3

R.T.: 9.026 min
Delta R.T.: 0.014 min
Response: 2340550
Conc: 32.97 ng/ml



#38 AR-1262-3


R.T.: 7.950 min
Delta R.T.: 0.001 min
Response: 959288
Conc: 21.44 ng/ml



#39 AR-1262-4


R.T.: 9.100 min
Delta R.T.: -0.007 min
Response: 1468864
Conc: 25.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#39 AR-1262-4

R.T.: 8.016 min
Delta R.T.: 0.000 min
Response: 2752843
Conc: 33.20 ng/ml




#40 AR-1262-5

R.T.: 9.812 min
Delta R.T.: 0.001 min
Response: 774119
Conc: 23.19 ng/ml



#40 AR-1262-5


R.T.: 8.542 min
Delta R.T.: 0.003 min
Response: 713876
Conc: 22.54 ng/ml



#41 AR-1268-1


R.T.: 9.026 min
Delta R.T.: 0.018 min
Response: 2340550
Conc: 18.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#41 AR-1268-1

R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 959288
Conc: 7.66 ng/ml




#42 AR-1268-2

R.T.: 9.100 min
Delta R.T.: -0.010 min
Response: 1468864
Conc: 13.32 ng/ml



#42 AR-1268-2


R.T.: 8.016 min
Delta R.T.: 0.000 min
Response: 2752843
Conc: 23.77 ng/ml



#43 AR-1268-3


R.T.: 9.357 min
Delta R.T.: 0.002 min
Response: 69169
Conc: 0.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.230 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.812 min
Delta R.T.: 0.002 min
Response: 774119
Conc: 25.41 ng/ml



#44 AR-1268-4


R.T.: 8.542 min
Delta R.T.: 0.004 min
Response: 713876
Conc: 21.10 ng/ml



#45 AR-1268-5

R.T.: 10.255 min
Delta R.T.: -0.002 min
Response: 270114
Conc: 0.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2024



#45 AR-1268-5

R.T.: 8.862 min
Delta R.T.: 0.002 min
Response: 211739
Conc: 0.73 ng/ml