

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082224\
 Data File : PP066814.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2024 21:28
 Operator : YP\AJ
 Sample : P3390-03
 Misc : AR1232 MDL 25 PPB
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MDL-SOIL-03-QT3-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 23:11:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 21 05:19:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.705	3.985	27640773	18456951	18.666	14.116
2) SA Decachlor...	10.625	9.144	16785329	13060977	19.127	15.678
Target Compounds						
11) L3 AR-1232-1	5.077	4.368	787849	566005	24.773	18.470 #
12) L3 AR-1232-2	5.610	5.109	466628	517955	26.940	19.196 #
13) L3 AR-1232-3	5.900	5.294	778149	183219	25.777	13.662 #
14) L3 AR-1232-4	6.065	5.372	326983	177884	22.201	13.838 #
15) L3 AR-1232-5	6.151	5.553	280230	222851	23.607	16.092 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082224\
Data File : PP066814.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2024 21:28
Operator : YP\AJ
Sample : P3390-03
Misc : AR1232 MDL 25 PPB
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
MDL-SOIL-03-QT3-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 23:11:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP082024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 21 05:19:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

