

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP082224\  
 Data File : PP066827.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 23 Aug 2024 01:22  
 Operator : YP\AJ  
 Sample : P3390-09  
 Misc : AR1254 MDL 25 PPB  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 MDL-WATER-03-QT3-2024

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 08/23/2024  
 Supervised By :Ankita Jodhani 08/23/2024

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 23 06:34:28 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP082024.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Aug 21 05:19:45 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml    |
|-----------------------------|--------|-------|----------|----------|---------|----------|
| -----                       |        |       |          |          |         |          |
| System Monitoring Compounds |        |       |          |          |         |          |
| 1) SA Tetrachlo...          | 4.706  | 3.986 | 23591942 | 15593155 | 15.932  | 11.926 # |
| 2) SA Decachlor...          | 10.623 | 9.144 | 14500651 | 11253071 | 16.524  | 13.508   |
| Target Compounds            |        |       |          |          |         |          |
| 26) L6 AR-1254-1            | 6.737  | 5.910 | 1276491  | 1146045  | 25.857  | 19.131 # |
| 27) L6 AR-1254-2            | 6.954  | 6.058 | 1692864  | 1124923  | 23.203m | 21.050   |
| 28) L6 AR-1254-3            | 7.324  | 6.471 | 1722725  | 1531785  | 24.656  | 19.372   |
| 29) L6 AR-1254-4            | 7.608  | 6.701 | 905276   | 680051   | 20.333  | 16.822m  |
| 30) L6 AR-1254-5            | 8.026  | 7.125 | 1430086  | 1184063  | 21.571m | 18.422   |
| -----                       |        |       |          |          |         |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP082224\  
Data File : PP066827.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 23 Aug 2024 01:22  
Operator : YP\AJ  
Sample : P3390-09  
Misc : AR1254 MDL 25 PPB  
ALS Vial : 33 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
MDL-WATER-03-QT3-2024

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 08/23/2024  
Supervised By :Ankita Jodhani 08/23/2024

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 23 06:34:28 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP082024.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Aug 21 05:19:45 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

