

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082321\
 Data File : PP038794.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Aug 2021 19:17
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 07:35:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 18 04:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.882	3.868	2013120	1299946	49.882	54.246
2)	SA Decachlor...	10.868	9.127	1346715	1287584	54.736	52.866
Target Compounds							
21)	L5 AR-1248-1	6.201	5.111	365949	261722	459.268	480.300
22)	L5 AR-1248-2	6.495	5.374	445778	356731	454.668	479.980
23)	L5 AR-1248-3	6.712	5.418	526656	370918	456.247	488.461
24)	L5 AR-1248-4	7.141	5.601	559748	443015	453.288	478.955
25)	L5 AR-1248-5	7.181	6.021	550021	456170	449.850	505.229



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082321\
Data File : PP038794.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Aug 2021 19:17
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 07:35:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP081721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 18 04:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.882 min
Delta R.T.: -0.006 min
Response: 2013120
Conc: 49.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.868 min
Delta R.T.: -0.006 min
Response: 1299946
Conc: 54.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.868 min
Delta R.T.: -0.009 min
Response: 1346715
Conc: 54.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.127 min
Delta R.T.: -0.009 min
Response: 1287584
Conc: 52.87 ng/ml



#21 AR-1248-1


R.T.: 6.201 min
Delta R.T.: -0.003 min
Response: 365949
Conc: 459.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.111 min
Delta R.T.: -0.004 min
Response: 261722
Conc: 480.30 ng/ml




#22 AR-1248-2

R.T.: 6.495 min
Delta R.T.: -0.003 min
Response: 445778
Conc: 454.67 ng/ml



#22 AR-1248-2


R.T.: 5.374 min
Delta R.T.: -0.004 min
Response: 356731
Conc: 479.98 ng/ml



#23 AR-1248-3


R.T.: 6.712 min
Delta R.T.: -0.003 min
Response: 526656
Conc: 456.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.418 min
Delta R.T.: -0.004 min
Response: 370918
Conc: 488.46 ng/ml




#24 AR-1248-4

R.T.: 7.141 min
Delta R.T.: -0.003 min
Response: 559748
Conc: 453.29 ng/ml



#24 AR-1248-4


R.T.: 5.601 min
Delta R.T.: -0.004 min
Response: 443015
Conc: 478.95 ng/ml



#25 AR-1248-5

R.T.: 7.181 min
Delta R.T.: -0.002 min
Response: 550021
Conc: 449.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.021 min
Delta R.T.: -0.004 min
Response: 456170
Conc: 505.23 ng/ml