

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082322\
 Data File : PP050752.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2022 12:34
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 02:20:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.586	86791105	168.3E6	53.919	53.729
2) SA Decachlor...	10.097	8.549	84300953	70295143	47.874	45.673
Target Compounds						
21) L5 AR-1248-1	5.524	4.657	20761941	30515853	514.198	456.589
22) L5 AR-1248-2	5.797	4.892	28720261	42619411	514.089	493.370
23) L5 AR-1248-3	6.003	4.933	32810070	44369344	512.901	489.430
24) L5 AR-1248-4	6.410	5.104	38422513	49688103	485.727	471.430
25) L5 AR-1248-5	6.448	5.494	36368497	49148616	479.074	495.007



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082322\
Data File : PP050752.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2022 12:34
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 02:20:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.347 min
Delta R.T.: 0.001 min
Response: 86791105
Conc: 53.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.586 min
Delta R.T.: -0.006 min
Response: 168289636
Conc: 53.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: 0.008 min
Response: 84300953
Conc: 47.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.549 min
Delta R.T.: -0.007 min
Response: 70295143
Conc: 45.67 ng/ml



#21 AR-1248-1


R.T.: 5.524 min
Delta R.T.: 0.003 min
Response: 20761941
Conc: 514.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.657 min
Delta R.T.: -0.007 min
Response: 30515853
Conc: 456.59 ng/ml




#22 AR-1248-2

R.T.: 5.797 min
Delta R.T.: 0.003 min
Response: 28720261
Conc: 514.09 ng/ml



#22 AR-1248-2


R.T.: 4.892 min
Delta R.T.: -0.007 min
Response: 42619411
Conc: 493.37 ng/ml



#23 AR-1248-3


R.T.: 6.003 min
Delta R.T.: 0.003 min
Response: 32810070
Conc: 512.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.933 min
Delta R.T.: -0.007 min
Response: 44369344
Conc: 489.43 ng/ml




#24 AR-1248-4

R.T.: 6.410 min
Delta R.T.: 0.004 min
Response: 38422513
Conc: 485.73 ng/ml



#24 AR-1248-4


R.T.: 5.104 min
Delta R.T.: -0.007 min
Response: 49688103
Conc: 471.43 ng/ml



#25 AR-1248-5

R.T.: 6.448 min
Delta R.T.: 0.004 min
Response: 36368497
Conc: 479.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.494 min
Delta R.T.: -0.007 min
Response: 49148616
Conc: 495.01 ng/ml