

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082322\
 Data File : PP050760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2022 14:52
 Operator : YP\AJ
 Sample : N4294-04 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 45

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/24/2022
 Supervised By :Ankita Jodhani 08/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 02:25:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 09 11:38:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.342	3.588	3156516	5449974	1.961	1.740
2) SA Decachlor...	10.088	8.550	74493	2304606	0.042m	1.497 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP082322\
Data File : PP050760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2022 14:52
Operator : YP\AJ
Sample : N4294-04 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
45


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/24/2022
Supervised By :Ankita Jodhani 08/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 02:25:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 09 11:38:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.342 min
Delta R.T.: -0.004 min
Response: 3156516
Conc: 1.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
45


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/24/2022
Supervised By :Ankita Jodhani 08/24/2022




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.005 min
Response: 5449974
Conc: 1.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.088 min
Delta R.T.: -0.002 min
Response: 74493
Conc: 0.04 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.550 min
Delta R.T.: -0.005 min
Response: 2304606
Conc: 1.50 ng/ml